

Product information



Distribuito in ITALIA da
Li StarFish S.r.l.
Via Cavour, 35
20063 Cernusco S/N (MI)
telefono 02-92150794
info@listarfish.it
www.listarfish.it

Casein ELISA

REF

DECASE01



96



Demeditec Diagnostics GmbH
Lise-Meitner-Strasse 2
24145 Kiel – Germany
www.demeditec.com

TABLE OF CONTENTS / INHALTSVERZEICHNIS

1. GENERAL INFORMATION	3
2. PRINCIPLE OF THE TEST	3
3. PRECAUTIONS.....	3
4. HEALTH AND SAFETY INSTRUCTIONS.....	3
5. REAGENTS.....	4
6. ADDITIONAL INSTRUMENTATION AND REAGENTS (NOT PROVIDED).....	4
7. SAMPLE PREPARATION	5
8. PROCEDURE.....	5
9. CALCULATION OF RESULTS.....	6
10. TYPICAL STANDARD VALUES	6
11. PERFORMANCE.....	7

SYMBOLS USED WITH DEMEDITEC ASSAYS	16
--	----

Sensitivity (Casein)	0.00-0.03 ppm
Recovery	82-102%
Incubation Time	60 min

1. GENERAL INFORMATION

Bovine milk belongs to the most important allergenic food ingredients especially for children. Already very low amounts of bovine milk can cause allergic reactions, which may lead to anaphylactic shock in severe cases. Because of this, milk allergic persons must strictly avoid the consumption of milk or milk containing food. In particular the presence of hidden milk proteins such as in sausage, cookies, convenience food or beverages represent a critical problem for milk allergic persons. According to EU Directive 2003/89/EG the addition of bovine milk has to be labeled. For the detection of bovine milk in foodstuffs, sensitive detection systems are required. Approximately 80% of bovine milk proteins are caseins which are composed of α -, β - and κ -caseins. So these heat-stable allergens represent the main fraction of bovine milk proteins.

The Demeditec Casein ELISA represents a highly sensitive detection system and is particularly capable of the identification and quantification of bovine casein residues in cookies, bread crumbs, sausage, orange juice, wine, soy products and chocolate.

2. PRINCIPLE OF THE TEST

The **Casein** quantitative test is based on the principle of the enzyme linked immunosorbent assay. An antibody directed against casein is bound on the surface of a microtiter plate. Casein containing samples or standards are given into the wells of the microtiter plate. After 20 minutes incubation at room temperature, the wells are washed with diluted washing solution to remove unbound material. A peroxidase conjugated second antibody directed against casein is given into the wells and after 20 minutes of incubation the plate is washed again. A substrate solution is added and incubated for 20 minutes, resulting in the development of a blue colour. The colour development is inhibited by the addition of a stop solution, and the colour turns yellow. The yellow colour is measured photometrically at 450 nm. The concentration of casein is directly proportional to the colour intensity of the test sample.

3. PRECAUTIONS

Full compliance of the following good laboratory practices (GLP) will determine the reliability of the results:

1. Prior to beginning the assay procedure, bring all reagents to room temperature (20-25°C).
2. All reagents should be mixed by gentle inversion or swirling prior to use. Do not induce foaming.
3. Once the assay has been started, all subsequent steps should be completed without interruption and within the recommended time limits.
4. Replace caps in all the reagents immediately after use. Do not interchange vial stoppers.
5. Use a separate disposable tip for each specimen to prevent cross-contamination.
6. All specimens and standards should be run at the same time, so that all conditions of testing are the same.
7. Do not mix components from different batches.
8. Do not use reagents after expiration date.
9. Check both precision and accuracy of the laboratory equipment used during the procedure (micropipets, ELISA reader etc.).

4. HEALTH AND SAFETY INSTRUCTIONS

1. Do not smoke or eat or drink or pipet by mouth in the laboratory.
2. Wear disposable gloves whenever handling patient specimens.
3. Avoid contact of substrate and stop solution with skin and mucosa (possible irritation, burn or toxicity hazard). In case of contact, rinse the affected zone with plenty of water.
4. Handling and disposal of chemical products must be done according to good laboratory practices (GLP).

5. REAGENTS

The kit contains reagents for 96 determinations. They have to be stored at 2-8°C. Expiry data are found on the labels of the bottles and the outer package.

1. **SORB MT** Microtiter plate consisting of 12 strips with 8 breakable wells each, coated with anti-casein antibodies.
2. **CAL 1 – 5 100x** Casein Standards (0, 0.2, 0.6, 2, 6 ppm of casein): 5 vials with 2.0 mL as 100x concentrate, dyed brownish. Dilute 20 µL of standard with 1980 µL **pre-diluted** extraction and sample dilution buffer to achieve the concentrations named above. Stored at 4°C the diluted standards are stable for at least 24 hours.
Note: The concentrations above refer to the 100x diluted standards.
3. **ENZ CONJ** Conjugate (anti-casein-peroxidase): 15 mL, dyed red, ready-to-use.
4. **SUB TMB** Substrate Solution (TMB): 15 mL, ready-to-use.
5. **STOP SOLN** Stop Solution (0.5 M H₂SO₄): 15 mL, ready-to-use.
6. **SAM DIL 5x** Extraction and sample dilution buffer (Carbonate buffer): 2 x 120 mL as 5x concentrate, dyed red. Dilute 1+4 with distilled water. Stored at 4°C the diluted buffer is stable for at least one week. If during the cold storage crystals precipitate, the concentrate should be warmed up to 37°C for 15 minutes.
7. **WASH SOLN 10x** Washing Solution (PBS + Tween 20): 60 mL as 10x concentrate. Dilute 1+9 with distilled water. Stored at 4°C the diluted buffer is stable for at least 4 weeks. If during the cold storage crystals precipitate, the concentrate should be warmed up to 37°C for 15 minutes.
8. Instruction Manual.

6. ADDITIONAL INSTRUMENTATION AND REAGENTS (NOT PROVIDED)

Instrumentation

- 100 - 1000 µL micropipets
- Analytical balance
- Mortar, mixer
- Water bath
- Centrifuge
- ELISA reader (450 nm)
- Plastic bag to store unused microtiter strips

Reagents

- double distilled water

7. SAMPLE PREPARATION

Due to a high risk of cross-contamination all applied instruments like applicator, mortar, glass vials etc. have to be **cleaned thoroughly** before and after each sample. To identify possible cross-contamination caused by previous extractions it is strongly recommended to note the sequence of the extractions.

The following sample preparation should be applied for solid samples:

1. To maximize homogeneity and representativeness of the sample drawing, a minimum of 5 g sample should be pulverized finely in a mortar, impact mill etc.
2. 0.5 g of the homogenized mixture is suspended in 10 mL of **pre-diluted** extraction and sample dilution buffer. Afterwards the suspension is incubated for 15 min in a preheated water bath at 60°C. To ensure good homogeneity, the samples should be shaken every two minutes.
3. The samples are centrifuged for 10 minutes at 2000 g. If it is not possible to separate the supernatant from the precipitate completely, the suspension should be filtrated if necessary.
4. Due to high matrix effects meat and sausage samples should be further diluted 1 + 4 with **pre-diluted** extraction and sample dilution buffer.
5. 100 µL of particle-free solution are applied per well. If the results of a sample are out of the measuring range, further dilution with the **pre-diluted** extraction and sample dilution buffer is necessary. The additional dilution has to be considered when calculating the concentration.

The following sample preparation should be applied for liquid samples:

0.5 mL of liquid sample is diluted in 9.5 mL of **pre-diluted** extraction and sample dilution buffer. Afterwards the suspension is incubated for 15 min in a preheated water bath at 60°C. To ensure good homogeneity, the samples should be shaken every two minutes. The process is continued at point 3 of solid sample extraction process.

8. PROCEDURE

The washing solution is supplied as 10x concentrate and has to be **diluted** 1+9 with double distilled water before use.

In any case the **diluted** standards should be determined at least twofold. When samples in great numbers are determined, the standards should be pipetted once before the samples and once after the samples. For final interpretation the arithmetic mean is used for calculation.

In consideration of GLP and quality control requirements a duplicate measurement of samples is recommended. The procedure is according to the following scheme:

1. Prepare samples as described above.
2. Pipet 100 µL of **diluted** standards or prepared samples in duplicate into the appropriate wells of the microtiter plate.
3. Incubate for 20 minutes at room temperature.
4. Wash the plate three times as follows: Discard the contents of the wells (dump or aspirate). Pipet 300 µL of diluted washing solution into each well. After the third repetition empty the wells again and remove residual liquid by striking the plate against a paper towel. The wash procedure is critical. Insufficient washing will result in poor precision and falsely elevated absorbances.
5. Pipet 100 µL of conjugate (anti-casein-peroxidase) into each well.
6. Incubate for 20 minutes at room temperature.
7. Wash the plate as outlined in 4.
8. Pipet 100 µL of substrate solution into each well.
9. Allow the reaction to develop in the dark (e.g. cupboard or drawer; the chromogen is light-sensitive) for 20 minutes at room temperature.
10. Stop enzyme reaction by adding 100 µL of stop solution (0.5 M H₂SO₄) into each well. The blue colour will turn yellow upon addition.
11. After thorough mixing, measure absorbance at 450 nm (reference wavelength 620 nm), using an ELISA reader. The colour is stable for 30 minutes.

9. CALCULATION OF RESULTS

The **diluted** standards are prepared for a direct determination of sample concentrations. The dilution of samples in the extraction process as described in the above stated sample preparation procedure is already considered. Additional dilution due to meat containing samples or high sample concentration has to be accounted for.

1. Calculate the average optical density (OD 450 nm) for each set of reference standards or samples.
2. Construct a standard curve by plotting the mean optical density obtained for each reference standard against its concentration in ppm on semi-log graph paper with the optical density on the vertical (y) axis and the concentration on the horizontal (x) axis. Alternatively the evaluation can be carried out by software. In this case the 4-parameter method should be preferred.
3. Using the mean optical density value for each sample, determine the corresponding concentration of casein in ppm from the standard curve. Depending on experience and/or the availability of computer capability, other methods of data reduction may be employed.

For calculation of the amount of a corresponding raw product, the casein concentration has to be multiplied with a product specific conversion factor (F).

The following conversion factors have been determined by means of validation experiments:

Whole milk	42
Skim milk powder (MoniQA MQA 092014)	4,4
Skim milk powder(NIST RM1549)	3,6
Whole Milk Powder (NIST RM8435)	4,9
Caseinate	1,2

10. TYPICAL STANDARD VALUES

The following table contains an example for a typical standard curve. The binding is calculated as percent of the absorption of the 6 ppm standard. These values are only an example and should not be used instead of the standard curve which has to be measured in each new test.

Casein (ppm)	% binding of 6 ppm
6	100
2	75
0.6	44
0.2	23
0	8

11. PERFORMANCE

Sensitivity

The limit of detection (LOD) of the **Demeditec Casein test** is 0.03 ppm casein.

Validation experiments with common matrices resulted in the following LODs [ppm]:

Cookies	0.02
Bread Crumbs	0.03
Sausage	0.02
Chocolate	0.00
White wine	0.02
Red Wine	0.02
Soy Milk	0.02
Orange Juice	0.03

The limit of quantification (LOQ) of the **Demeditec Casein test** is 0.2 ppm casein.

Due to the variety of sample matrices and their influence on the blank, results less than the LOQ should be treated as negative.

Specificity

For the following foods no cross-reactivity could be detected:

Almond	Hazelnut	Pumpkin seed
Barley	Lecithin	Rice, brown
Beef	Lima bean	Rice, white
Brazil nut	Oats	Rye
Buckwheat	Pea	Sacharose
Chicken	Peanut	Sesame
Chickpea	Pecan nut	Shrimps
Cocoa	Pineseed	Soy
Coconut	Pistachio	Split peas
Cod	Poppy seed	β -Lactoglobulin
Corn	Pork	Walnut
Egg	Prawn	Wheat

The following cross-reactions were determined:

Ewe's milk	< 1.2%
Goat's milk	< 1.1%

Precision

Intra-assay Precision	5 - 11%
Inter-assay Precision	8 - 14%

Linearity














The serial dilution of spiked samples (cookies, bread crumbs, chocolate, sausage, soy milk, orange juice and white wine) resulted in a dilution linearity of 100% - 127%.

Recovery

Mean recovery was determined by spiking samples with different amounts of casein:

Cookies	99%
Bread crumbs	84%
Chocolate	85%
Sausage	82%
Soy milk	89%
Orange juice	87%
White wine	102%
Red wine	102%

SYMBOLS USED WITH DEMEDITEC ASSAYS

Symbol	English	Deutsch	Française	Espanol	Italiano
	European Conformity	CE-Konformitätskennzeichnung	Conforme aux normes européennes	Conformidad europea	Conformità europea
	Consult instructions for use	Gebrauchsanweisung beachten	Consulter les instructions d'utilisation	Consulte las Instrucciones	Consultare le istruzioni per l'uso
	In vitro diagnostic device	In-vitro-Diagnostikum	utilisation Diagnostic in vitro	Diagnóstico in vitro	Per uso Diagnostica in vitro
	For research use only	Nur für Forschungszwecke	Seulement dans le cadre de recherches	Sólo para uso en investigación	Solo a scopo di ricerca
	Catalogue number	Katalog-Nr.	Référence	Número de catálogo	No. di catalogo
	Lot. No. / Batch code	Chargen-Nr.	No. de lot	Número de lote	Lotto no
	Contains sufficient for <n> tests/	Ausreichend für "n" Ansätze	Contenu suffisant pour "n" tests	Contenido suficiente para <n> ensayos	Contenuto sufficiente per "n" saggi
	Note warnings and precautions	Warnhinweise und Vorsichtsmaßnahmen beachten	Avertissements et mesures de précaution font attention	Tiene en cuenta advertencias y precauciones	Annoti avvisi e le precauzioni
	Storage Temperature	Lagerungstemperatur	Température de conservation	Temperatura de conservación	Temperatura di conservazione
	Expiration Date	Mindesthaltbarkeitsdatum	Date limite d'utilisation	Fecha de caducidad	Data di scadenza
	Legal Manufacturer	Hersteller	Fabricant	Fabricante	Fabbricante
<i>Distributed by</i>	Distributed by	Vertrieb durch	Distribution par	Distribución por	Distribuzione da parte di
	Version	Version	Version	Versión	Versione
	Single-use	Einmalverwendung	À usage unique	Uso único	Uso una volta