



Distribuito in ITALIA da
Li StarFish S.r.l.
Via Cavour, 35
20063 Cernusco S/N (MI)
telefono 02-92150794
fax 02-92157285
info@listarfish.it
www.listarfish.it

Human IFN- γ ELISpot^{PLUS}

Product Code: 3420-4AST-2

CONTENTS:

2 pre-coated strip plates, mAb 1-D1K
Empty plate frame included for transfer of strips

Vial 1 (yellow top)

Biotinylated detection mAb 7-B6-1 (40 μ l)
Concentration 1 mg/ml

Vial 2 (white top)

Streptavidin-ALP (40 μ l)

Vial 3 (black top)

Positive control anti-CD3 mAb CD3-2 (100 μ l)

BCIP/NBT-plus substrate (25 ml)

The detection antibody is supplied in sterile filtered (0.2 μ m) PBS with 0.02% sodium azide. Streptavidin-ALP is supplied in 0.1 M Tris buffer with 0.002% Kathon CG. Anti-CD3 mAb is supplied in sterile filtered (0.2 μ m) PBS. Vials have been overfilled to ensure recovery of the specified amount.

STORAGE:

Shipped at ambient temperature. On arrival all reagents should be stored refrigerated at 4-8°C. Plates should be kept at room temperature.

Guidelines for Human IFN- γ ELISpot^{PLUS}

Please read through before starting the assay

A Preparation of ELISpot plate (sterile conditions)

1. Assemble the required number of strips in the extra plate frame and wash 4 times with sterile PBS (200 μ l/well). Seal the bag with the remaining strips and store at room temperature.
2. Condition the plate with medium (200 μ l/well) containing 10% of the same serum as used for the cell suspensions. Incubate for at least 30 minutes at room temperature.

B Incubation of cells in plate (sterile conditions)

1. Remove the medium and add the stimuli followed by the cell suspension. Alternatively cells and stimuli can be mixed before addition to the plate. The mAb CD3-2, included in the kit, is recommended as a positive control for cytokine production at a dilution of 1:1000.
2. Put the plate in a 37°C humidified incubator with 5% CO₂ and incubate for 12-48 hours. Do not move the plate during this time and take measures to avoid evaporation (e.g. by wrapping the plate in aluminium foil).

C Detection of spots

1. Remove the cells by emptying the plate and wash 5 times with PBS, 200 μ l/well.
2. Dilute the detection antibody (7-B6-1-biotin) to 1 μ g/ml in PBS containing 0.5% fetal calf serum (PBS-0.5% FCS). Add 100 μ l/well and incubate for 2 hours at room temperature.
3. Wash plate as above (step C1).
4. Dilute the Streptavidin-ALP (1:1000) in PBS-0.5% FCS and add 100 μ l/well. Incubate for 1 hour at room temperature.
5. Wash plate as above (step C1).
6. Filter the ready-to-use substrate solution (BCIP/NBT-plus) through a 0.45 μ m filter and add 100 μ l/well. Develop until distinct spots emerge.
7. Stop color development by washing extensively in tap water. Remove the plate frame from the plastic tray and rinse the underside of the membrane.
8. Leave the plate to dry. Inspect and count spots in an ELISpot reader or in a dissection microscope.
9. Store plate in the dark at room temperature.

Hints and Comments

These suggestions are based on the detection of antigen-specific immune responses using PBMC. If using clones, mixtures of separated cell fractions etc., other protocols may have to be considered.

Plate washing

Washing of plates can be done using a multi-channel micropipette. In washing steps not requiring sterile conditions (C1-C5), a regular ELISA plate washer can also be used, provided that the washing head is adapted to the ELISpot plates.

Cells

Both freshly prepared and cryopreserved cells may be used in the assay. However it is recommended that the latter are rested for at least one hour to allow removal of cell debris before addition to the plate. Triplicates or duplicates of 250,000 cells per well are often used to assess antigen-specific responses. For polyclonal activators, the cell number may have to be reduced to avoid confluent spot formation. Protocols with other incubation times have to be established by the user.

Serum

The serum should be selected to support cell culture and give low background staining. We recommend the use of fetal calf serum. Alternatively serum-free medium evaluated for cell culture can be used. Human serum is not recommended as it may contain heterophilic antibodies or intrinsic analyte which may interfere with the assay.

Assay controls

The number of cells responding to stimulation is often compared to the number of cells spontaneously producing the cytokine, which is determined by incubating the same number of cells in the absence of stimuli. A polyclonal activator, such as the included anti-CD3 mAb or phytohemagglutinin (1-10 $\mu\text{g}/\text{ml}$), is often used as a control for cell viability and functionality of the test system.

Detection antibody

To reduce unspecific background it is recommended to filter (0.2 μm) the working dilution of detection mAb.

Buffers

PBS for washing and dilution should be filtered (0.2 μm) for optimal results. Avoid the inclusion of Tween or other detergents in the washing and incubation buffers.

Substrate development

Development is made until distinct spots are visible in positive wells (usually 10-30 minutes). A general darkening of the membrane may occur during development but disappears after drying.



Distribuito in ITALIA da
Li StarFish S.r.l.
Via Cavour, 35
20063 Cernusco S/N (MI)
telefono 02-92150794
fax 02-92157285
info@listarfish.it
www.listarfish.it

MABTECH

NOTE; for research use only.

MABTECH shall not be liable for the use or handling of the product or for consequential, special, indirect or incidental damages therefrom.



Intertek



Intertek

Developed and manufactured by MABTECH AB, Sweden, whose quality management system complies with the above standards.

Notice to purchasers: Limited Licence

Licences under International Patent Application PCT/GB97/03222, other applications based on that Application, and patents issuing in response (the Patents), owned by Isis Innovation Limited, Oxford, UK (Isis), and licences under other intellectual property rights owned by Isis and related to the patented technology, are restricted by reference to the use which is made of the patented technology. The purchase price of this product includes limited, non-transferable rights to practise the method described in the Patents solely in connection with the activities of the purchaser in the diagnosis measurement, monitoring and assessment of the activity of T-cells in body fluids for research purposes only. No right to conduct any other activities by means of the method described in the Patents is granted by implication or estoppel.

Updated on 2015-02-24