

# EDI™ Novel Coronavirus COVID-19 IgG ELISA Kit

Enzyme Linked Immunosorbent Assay (ELISA) for the qualitative detection of the COVID-19 IgG in human serum.

REF:KT-1032 IVD CE     

## INTENDED USE

This kit is for in vitro diagnostics use only. The kit is detecting novel COVID-19 IgG antibody in human serum. It is for screening or to aid in the diagnosis of COVID-19. Patients with suspected clustering cases require diagnosis or differential diagnosis of novel coronavirus infection. The assay is validated manually, but can be adapted to an automated instrument. The assay is for the qualitative detection only.

## INTENDED USER

This kit is for laboratory professional use or healthcare professionals.

## SUMMARY OF PHYSIOLOGY

2019 novel coronavirus (COVID-19) is a single-stranded RNA coronavirus<sup>2</sup>. Comparisons of the genetic sequences of this virus have shown similarities to SARS-CoV and bat coronaviruses<sup>7</sup>. In humans, coronaviruses cause respiratory infections<sup>3</sup>. Coronaviruses are composed of several proteins including the spike (S), envelope (E), membrane (M), and nucleocapsid (N)<sup>4</sup>. Results suggest that the spike protein retains sufficient affinity to the Angiotensin converting enzyme-2 (ACE-2) receptor to use it as a mechanism of cell entry<sup>6</sup>. Human to human transmission of coronaviruses is primarily thought to occur among close contacts via respiratory droplets generated by sneezing and coughing<sup>1</sup>. IgG is the most abundantly found immunoglobulin to be produced in response to an antigen and will be maintained in the body after initial exposure for long term response<sup>5</sup>.

## ASSAY PRINCIPLE

This ELISA kit is designed, developed, and produced for the qualitative measurement of the human anti-COVID-19 IgG antibody in serum. This assay utilizes the microplate based enzyme immunoassay technique.

Assay controls and 1:100 diluted human serum samples are added to the microtiter wells of a microplate that was coated with COVID-19 recombinant full length nucleocapsid protein. After the first incubation period, the unbound protein matrix is removed with a subsequent washing step. A horseradish peroxidase (HRP) labeled polyclonal goat anti-human IgG tracer antibody is added to each well. After an incubation period, an immunocomplex of "COVID-19 recombinant antigen – human anti-COVID-19 IgG antibody - HRP labeled anti-human IgG tracer antibody" is formed if there is specific coronavirus IgG antibody present in the tested specimen. The unbound tracer antibody is removed by the subsequent washing step. HRP-labeled tracer antibody bound to the well is then incubated with a substrate solution in a timed reaction and then measured in a spectrophotometric microplate reader. The enzymatic activity of the tracer antibody bound to the anti-COVID-19 IgG on the wall of the microtiter well is proportional to the amount of the anti-COVID-19 IgG antibody level in the tested specimen.

## REAGENTS: PREPARATION AND STORAGE

This test kit must be stored at 2 – 8°C upon receipt. For the expiration date of the kit refer to the label on the kit box. All components are stable until this expiration date.

### **1. COVID-19 antigen coated Microplate (31217)**

Microplate coated with COVID-19 recombinant protein.

Qty: 1 x 96 well microplate  
Storage: 2 – 8°C  
Preparation: Ready to use.

### **2. COVID-19 IgG Sample Diluent (31218)**

A ready-to-use sample dilution buffer.

Qty: 1 x 120 mL  
Storage: 2 – 8°C  
Preparation: Ready to use.

### **3. HRP labeled Anti-hlgG Tracer Antibody (31273)**

Concentrated (21x) HRP labeled polyclonal goat anti-human IgG antibody in a stabilized protein matrix.

Qty: 1 x 0.6 mL  
Storage: 2 – 8°C  
Preparation: 1:21 dilution before use.

### **4. Tracer Antibody Diluent(31274)**

A buffer matrix for diluting the HRP Labeled Anti-hlgG Tracer Antibody.

Qty: 1 x 12 mL  
Storage: 2 – 8°C  
Preparation: Mix with the 21x HRP labeled anti-hlgG tracer antibody before assay.

### **5. ELISA Wash Concentrate (10010)**

Surfactant in a phosphate buffered saline with non-azide preservative.

Qty: 1 x 30 mL  
Storage: 2 – 25°C  
Preparation: 30x Concentrate. The contents must be diluted with 870 mL distilled water and mixed well before use.

### **6. ELISA HRP Substrate (10020)**

Tetramethylbenzidine (TMB) with stabilized hydrogen peroxide.

Qty: 1 x 15 mL  
Storage: 2 – 8°C  
Preparation: Ready to use.

### **7. ELISA Stop Solution (10030)**

0.5 M sulfuric acid.

Qty: 1 x 15 mL  
Storage: 2 – 25°C  
Preparation: Ready to use.

### **8. COVID-19 IgG Negative Control (31221)**

Negative control with a bovine serum albumin based matrix with non-azide preservative. Control products do not contain any serum from patients with new type of coronavirus infection.

Qty: 1 x 1 mL  
Storage: 2 – 8°C.  
Preparation: Ready to use.

### **9. COVID-19 IgG Positive Control (31222)**

Positive control with a bovine serum albumin based matrix with non-azide preservative. Control products do not contain any serum from patients with new type of coronavirus infection.

Qty: 1 x 0.5 mL  
Storage: 2 – 8°C.  
Preparation: Ready to use.

## **SAFETY PRECAUTIONS**

The reagents are for in-vitro diagnostic use only. Source material which contains reagents of bovine serum albumin was derived in the contiguous 48 United States. It was obtained only from healthy donor animals maintained under veterinary supervision and found free of contagious diseases. Wear gloves while performing this assay and handle these reagents as if they were potentially infectious. Avoid contact with reagents containing hydrogen peroxide, or sulfuric acid. Do not get in eyes, on skin, or on clothing. Do not ingest or inhale fumes. On contact, flush with copious amounts of water for at least 15 minutes. Exercise Good Laboratory Practices.

## **MATERIALS REQUIRED BUT NOT PROVIDED**

1. Precision single channel pipettes capable of delivering 10  $\mu$ L, 25  $\mu$ L, 100  $\mu$ L, and 1000  $\mu$ L, etc.
2. Repeating dispenser suitable for delivering 100  $\mu$ L.
3. Disposable pipette tips suitable for above volume dispensing.
4. Disposable 12 x 75 mm or 13 x 100 mm plastic tubes.
5. Disposable plastic 1000 mL bottle with caps.
6. Aluminum foil.
7. Deionized or distilled water.
8. Plastic microtiter well cover or polyethylene film.
9. ELISA multichannels wash bottle or automatic (semi-automatic) washing system.
10. Spectrophotometric microplate reader capable of reading absorbance at 450 nm.

## **SAMPLE COLLECTION & STORAGE**

Only 10  $\mu$ L of human serum is required for measurement in duplicate. Samples should only be used on the same day. Severe hemolytic samples should not be used.

## **ASSAY PROCEDURE**

### **1. Reagent Preparation**

1. Prior to use, allow all reagents to come to room temperature. Reagents from different kit lot numbers should not be combined or interchanged.
2. ELISA Wash Concentrate (10010) must be diluted to working solution prior to use. Please see REAGENTS section for details.
3. Dilute the HRP-labeled anti-hlgG tracer antibody (31273) 1:21 with Tracer Antibody Diluent (31274) by pipetting **0.6mL** tracer antibody to the diluent bottle.

### **2. Sample Preparation**

1. Dilute sample by a 1:100 dilution ratio with the COVID-19 IgG Sample Diluent (31218). For each 10  $\mu$ L of sample, 1000  $\mu$ L of COVID-19 IgG Sample Diluent (31218) is needed.
2. Mix well prior to performing the assay.

### **3. Assay Procedure**

1. Place a sufficient number of microwell strips (31217) in a holder to run the negative control (31221) in triplicate, the positive control (31222) in singlet, and samples in duplicate.
2. Test Configuration

Row	Strip 1	Strip 2	Strip 3
A	Negative Control	SAMPLE 3	SAMPLE 7
B	Negative Control	SAMPLE 3	SAMPLE 7
C	Negative Control	SAMPLE 4	SAMPLE 8
D	Positive Control	SAMPLE 4	SAMPLE 8
E	SAMPLE 1	SAMPLE 5	SAMPLE 9
F	SAMPLE 1	SAMPLE 5	SAMPLE 9
G	SAMPLE 2	SAMPLE 6	SAMPLE 10
H	SAMPLE 2	SAMPLE 6	SAMPLE 10

3. Add **100  $\mu$ L** of controls (31221, 31222) and 1:100 diluted samples into the designated microwells.
4. Mix gently and cover the plate with one plate sealer and aluminum foil. Incubate at **room temperature (20-25  $^{\circ}$ C)** for **30 minutes**.
5. Remove the plate sealer. Aspirate the contents of each well. Wash each well **5 times** by dispensing **350  $\mu$ L** of diluted wash solution (10010) into each well, and then completely aspirate the contents. Alternatively, an automated microplate washer can be used.
6. Add **100  $\mu$ L** of the 1:21 diluted HRP labeled Anti-hlgG Tracer Antibody into the microwells.
7. Mix gently and cover the plate with one plate sealer and aluminum foil. Incubate at **room temperature (20-25  $^{\circ}$ C)** for **30 minutes**.
8. Remove the plate sealer. Aspirate the contents of each well. Wash each well **5 times** by dispensing **350  $\mu$ L** of diluted wash solution (10010) into each well, and then completely aspirate the contents. Alternatively, an automated microplate washer can be used.
9. Add **100  $\mu$ L** of the substrate (10020) into the microwells.
10. Mix gently and cover the plate with aluminum foil. Incubate at **room temperature (20-25  $^{\circ}$ C)** for **20 minutes**.
11. Remove the aluminum foil and add **100  $\mu$ L** of stop solution (10030) into each of the microwells. Mix by gently by tapping the plate.
12. Read the absorbance at **450 nm** within **10 minutes** with a microplate reader.

## **PROCEDURAL NOTES**

1. It is recommended that all samples be assayed in duplicate. The average absorbance reading of each duplicate should be used for data reduction and the calculation of results.
2. Keep light-sensitive reagents in the original bottles and avoid unnecessary exposure to the light.
3. Store any unused antibody-coated strips in the foil Ziploc bag with desiccant to protect from moisture.
4. Careful technique and use of properly calibrated pipetting devices are necessary to ensure reproducibility of the test.
5. Incubation times or temperatures other than those stated in this insert may affect the results.
6. Avoid air bubbles in the microwell as this could result in lower binding efficiency and higher CV% of duplicate reading.
7. All reagents should be mixed gently and thoroughly prior to use. Avoid foaming.

## **QUALITY CONTROL**

To assure the validity of the results each assay must include both negative and positive controls. The average value of the absorbance of the negative control is less than 0.25, and the absorbance of the positive control is not less than 0.30. We also recommend that all assays include the laboratory's own controls in addition to those provided with this kit.

## **INTERPRETION OF RESULTS**

1. Calculate the average value of the absorbance of the negative control (xNC).
2. Calculate the cutoffs using the following formulas:
  - Positive cut off =  $1.1 \times (xNC + 0.18)$
  - Negative cut off =  $0.9 \times (xNC + 0.18)$
3. Determine the interpretation of the sample by comparing the OD to the following table:

Interpretation	Interval	Results
Negative	Measured value $\leq$ negative cut off	The sample does not contain the new coronavirus (COVID-19) IgG-related antibody
Positive	Measured value $\geq$ positive cut off	The sample contains novel coronavirus (COVID-19) an IgG-associated antibodies.
Borderline	Negative cutoff < Measured value < Positive cut off	Retest the sample in conjunction with other clinical tests.

- Race and geographical region may affect the results from normal donor samples. Laboratories may establish or modify the cutoff based upon additional validation.

ID	Average OD (450 nm)	Results	CV (%)
Sample 1	1.11	12/12 are Positive	1.96%
Sample 2	0.65	12/12 are Positive	3.47%
Sample 3	0.19	12/12 are Negative	4.66%
Negative Control	0.17	12/12 are Negative	3.15%
Positive Control	0.65	12/12 are Positive	4.14%

## LIMITATIONS OF THE PROCEDURE

- This test is only for qualitative detection. Test results should not be the sole basis for clinical diagnosis and treatment. The confirmation of infection with novel coronavirus (COVID-19) must be combined with the patient's clinical signs in conjunction to other tests.
- In the first week of the onset of the infection with the novel coronavirus (COVID-19) patients results may be negative for IgG. In addition, patients with low immunity or other diseases that affect immune function, failure of important systemic organs, and use of drugs that suppress immune function can also lead to negative results of new coronavirus IgG. Previous infection of SARS or other coronavirus strain may cause a light IgG positive in view of similarity of different strains.
- Bacterial or fungal contamination of serum specimens or reagents, or cross-contamination between reagents may cause erroneous results.
- Water deionized with polyester resins may inactivate the horseradish peroxidase enzyme.

## PERFORMANCE CHARACTERISTICS

### Assay Development

This assay was developed by evaluating eight commercially available COVID-19 antigens to screen for optimal use in this serological test. The assays were first evaluated with normal healthy donor serum samples to obtain negative test results. The assays were further evaluated with 20 positive serum samples from confirmed COVID-19 patients tested by RT-PCR. The best performing antigen was selected for the development of the kit. During this time, cross-reactivity was identified, but eliminated during the final selection of the antigen.

### Reactivity/Inclusivity

Although mutations in the SARS-CoV-2 genome have been identified as the virus has spread, no serologically unique strains have been described relative to the originally isolated virus (this research is exceptionally limited at present).

### Limit of Detection

Three lots of material were tested with one assay using a blank control in sixteen replicates. The LoD was calculated as the mean of the OD for the blank control plus three times the standard deviation. The highest value of the three runs was established for the LoD at 0.0666. The results are as follows:

Run	Average OD (450 nm)	CV (%)	LOD ( $\bar{x} + 3 SD$ )
Run 1	0.0481	4.83%	0.0550
Run 2	0.0518	5.71%	0.0606
Run 3	0.0531	8.44%	0.0666

### Repeatability

One lot of material was tested with one assay using three samples (strong positive, light positive, and negative) in sixteen replicates. For all sixteen replicates, sample 1 and 2 are positive and in sample 3 is all negative. The repeatability results are very satisfactory with acceptable CV.

The results are as follows:

ID	Average OD (450 nm)	Results	CV (%)
Sample 1	1.071	16/16 are Positive	6.35%
Sample 2	0.631	16/16 are Positive	3.11%
Sample 3	0.199	16/16 are Negative	4.99%

### Reproducibility

One lot of material was tested over twelve assays using three samples (strong positive, light positive, and negative) in two replicates and a set of positive and negative controls in three replicates. For all twelve assays, sample 1 and 2 are positive and sample 3 is all negative. The results for reproducibility are very satisfactory with an acceptable CV.

The results are as follows:

### Class Specificity

To evaluate class specificity, Ten RT-PCR confirmed COVID-19 patient serum samples were tested in duplicate in the Epitope Diagnostics, Inc. IgG and IgM ELISA Kits. Samples 1 - 5 are IgM positive and IgG negative. Sample 1 is a natural and untreated IgM positive, IgG negative. Sample 2 - 5 were originally positive for IgG and IgM but used protein A/ProSep A to remove the IgG. Samples 6 - 10 are IgG positive and IgM negative. All samples 6 - 10 are natural and untreated IgG positive, IgM negative. There is 100% agreement between the results of this test. This demonstrates that the assay is specific to the detection of IgG class without cross reaction to COVID-19 IgM class. The results are as follows:

Sample ID	IgM Result	IgG Result
Sample 1	+	-
Sample 2	+	-
Sample 3	+	-
Sample 4	+	-
Sample 5	+	-
Sample 6	-	+
Sample 7	-	+
Sample 8	-	+
Sample 9	-	+
Sample 10	-	+

### Cross-Reactivity

A large number of known negative samples (N=570) collected in the US prior to December 2019 were tested from a population with a high prevalence of vaccination against, and/or infection with, the following viruses, and specificity of 99.8% is observed, cross-reactivity testing for the following viruses would not be expected at this time:

anti-influenza A (IgG and IgM)
anti-influenza B (IgG and IgM)
anti-HCV (IgG and IgM)
anti-HBV (IgG and IgM)
anti-Haemophilus influenzae (IgG and IgM)
anti-229E (alpha coronavirus)
anti-NL63 (alpha coronavirus)
anti-OC43 (beta coronavirus)
anti-HKU1 (beta coronavirus)
ANA
anti-respiratory syncytial virus (IgG and IgM)
anti-HIV

To demonstrate cross-reactivity of the test, Epitope Diagnostics, Inc. used the FDA required minimum of 5 individual samples tested in duplicate for each disease/infectious agent using natural specimen confirmed with commercially available diagnostic tests. All samples were sourced from natural specimens using sera from patients with the underlying diseases in the acute or convalescent stages of infection. The disease and Infection agents were selected based on recommendations from the FDA EUA Program. The recommendation also included Anti-haemophilus influenzae and rhinovirus, but this material was unable to be tested due to lack of availability.

Agent	Disease State Confirmation Test	Results
Influenza A	Viron/Serion	5/5, Negative
Influenza B	Viron/Serion	5/5, Negative
Respiratory syncytial virus	EIA, Viron/Serion	5/5, Negative
Hepatitis C Virus	Roche Ampliprep/Taqman	5/5, Negative
Antinuclear Antibodies	Bio-Rad Hep 2	5/5, Negative
Hepatitis B Virus	Siemens	5/5, Negative

## Transportation Stability

One lot of material was shipped from Epitope Diagnostics, Inc. in San Diego, CA to an external site in the United States and returned. The kit was packaged in a foam box with blue ice which was not changed for the duration of the study to simulate transport conditions. The kits were in this condition for a total of 12 days. A comparison of the values obtained before and after shipment demonstrates the stability of the materials. The results are as follows:

Before Shipment			After Shipment		
Well ID	OD	Average	Well ID	OD	Average
Negative	0.145	0.142	Negative	0.110	0.110
	0.146			0.113	
	0.134			0.107	
Positive	0.588	N/A	Positive	0.526	N/A
Neg. Cut-Off	0.290		Neg. Cut-Off	0.261	
Pos. Cut Off	0.354		Pos. Cut Off	0.319	

## CLINICAL TESTING

Patient samples were tested using the IgG ELISA kit at four sites: Center for Disease Control and Prevention in China, a University Hospital in China, a laboratory in the United States, and a University Hospital Laboratory in the United States. The combined cohort consisted of normal healthy patients with samples collected prior to the COVID-19 outbreak (N=624) and RT-PCR confirmed positive patients (N=187). The results are as follows:

		Confirmed Positive	Confirmed Negative
EDI™ Novel Coronavirus COVID-19 IgG ELISA Kit	Positive	184	1
	Negative	3	623
	Total	187	624
PPA	98.4%	95% CI (Wilson's Score):	0.954- 0.995
NPA	99.8%	95% CI (Wilson's Score):	0.991-0.9997

## WARRANTY

This product is warranted to perform as described in its labeling and literature when used in accordance with all instructions. Epitope Diagnostics, Inc. DISCLAIMS ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE, and in no event shall Epitope Diagnostics, Inc. be liable for consequential damages. Replacement of the product or refund of the purchase price is the exclusive remedy for the purchaser. This warranty gives you specific legal rights and you may have other rights, which vary from state to state.


## REFERENCES

- CDC (2020). Transmission of Novel Coronavirus (COVID-19).
- Chenja Yuan , Shi Jinsong , Qiudong An , Liu Chang , Li Xin , Qiang , Ruanji Shou , mountains . Wuhan 2019 Bioinformatics coronavirus genome analysis [J / OL]. Bioinformatics: 1-10 [2020-02-10].
- Fehr, A. R., & Perlman, S. (2015). Coronaviruses: An Overview of Their Replication and Pathogenesis. *Coronaviruses Methods in Molecular Biology*, 1–23. doi: 10.1007/978-1-4939-2438-7\_1
- Li, F., Li, W., Farzan, M., & Harrison, S. (2005). Structure of SARS coronavirus spike receptor-binding domain complexed with its receptor. doi: 10.2210/pdb2ajf/pdb
- Wu, L.-P., Wang, N.-C., Chang, Y.-H., Tian, X.-Y., Na, D.-Y., Zhang, L.-Y., Liang, G.-D. (2007). Duration of Antibody Responses after Severe Acute Respiratory Syndrome. *Emerging Infectious Diseases*, 13(10), 1562–1564. doi: 10.3201/eid1310.070576
- Xu, X., Chen, P., Wang, J., Feng, J., Zhou, H., Li, X., Hao, P. (2020). Evolution of the novel coronavirus from the ongoing Wuhan outbreak and modeling of its spike protein for risk of human transmission. *Science China Life Sciences*. doi: 10.1007/s11427-020-1637-5
- Zhou, P., Yang, X.-L., Wang, X.-G., Hu, B., Zhang, L., Zhang, W., Shi, Z.-L. (2020). A pneumonia outbreak associated with a new coronavirus of probable bat origin. *Nature*. doi: 10.1038/s41586-020-2012-7

## TECHNICAL ASSISTANCE AND CUSTOMER SERVICE

For technical assistance or place an order, please contact Epitope Diagnostics, Inc. at (858) 693-7877 or fax to (858) 693-7678.

This product is manufactured by

 **Epitope Diagnostics, Inc.**  
7110 Carroll Road  
San Diego, CA 92121, US

EC	REP	MDSS GmbH Schiffgraben 41, 30175 Hannover, Germany
----	-----	--

## GLOSSARY OF SYMBOLS (EN 980/ISO 15223)

 In Vitro Diagnostic Device

 European Conformity

 Lot Number

 Catalog Number

 Read instructions before use


 Number of Tests

 Store at

 Use by

 Keep away from heat and direct sun light

 Manufacturer

 Authorized Representative in Europe



Distribuito in ITALIA da  
**Li StarFish S.r.l.**  
Via Cavour, 35  
20063 Cernusco S/N (MI)  
telefono 02-92150794  
info@listarfish.it  
www.listarfish.it