

Certificate of Analysis

Product: Human Total GLP-1 ELISA Kit

REF KT-602

LOT N343


EXP 2020-11

Components/Reagents

	LOT	EXP
Standard 1	M651	2021 - 11
Standard 2	M652	2021 - 11
Standard 3	M653	2021 - 11
Standard 4	M654	2021 - 11
Standard 5	M655	2021 - 11
Control I	M656	2021 - 11
Control II	M657	2021 - 11
Total GLP-1 Tracer Antibody	N187	2022 - 01
Total GLP-1 Capture Antibody	M915	2022 - 01
Tracer Antibody Diluent	N273	2021 - 08
Streptavidin Coated Plate	M900	2021 - 07
Wash Concentrate (30x)	M990	2022 - 02
HRP Substrate	E079	2020 - 11
Stop Solution	M360	2021 - 08

QC Results

Standards	Concentration (pM)	Average OD450/650nm	CV%
1	0	0.010	5.0
2	3.1	0.071	2.0
3	6.6	0.178	1.4
4	18.7	0.587	1.2
5	60.5	2.799	3.7
Controls	Target (pM)	QC Test Value (pM)	
I	3.8- 5.9	4.7	4.7
II	9.5 - 14.8	10.9	5.4



The presented components/reagents have been tested and approved at EPITOPE Diagnostics, Inc.

ACCEPTED

QC Released by: _____

Date: May 10, 2019