



Distribuito in ITALIA da
Li StarFish S.r.l.
Via Cavour, 35
20063 Cernusco S/N (MI)
telefono 02-92150794
info@listarfish.it
www.listarfish.it


Manual

β_2 -Microglobulin ELISA

*For the in vitro determination of β_2 -microglobulin
in plasma, serum and urine*

Valid from 2019-03-28

REF K 6210



IVD **CE**



Immundiagnostik AG, Stubenwald-Allee 8a, 64625 Bensheim, Germany

Tel.: +49 6251 70190-0

Fax: + 49 6251 70190-363

e.mail: info@immundiagnostik.com

www.immundiagnostik.com

Table of Contents

1. INTENDED USE	15
2. INTRODUCTION	15
3. MATERIAL SUPPLIED	15
4. MATERIAL REQUIRED BUT NOT SUPPLIED	16
5. STORAGE AND PREPARATION OF REAGENTS	16
6. STORAGE AND PREPARATION OF SAMPLES	17
<i>Sample storage</i>	17
<i>Dilution of samples</i>	17
7. ASSAY PROCEDURE	17
<i>Principle of the test</i>	17
<i>Test procedure</i>	18
8. RESULTS	19
9. LIMITATIONS	20
10. QUALITY CONTROL	20
<i>Reference range</i>	20
11. PERFORMANCE CHARACTERISTICS	20
<i>Spiking Recovery</i>	20
<i>Dilution recovery</i>	20
<i>Precision and reproducibility</i>	21
<i>Analytical Sensitivity</i>	21
12. PRECAUTIONS	21
13. TECHNICAL HINTS	22
14. GENERAL NOTES ON THE TEST AND TEST PROCEDURE	22

1. INTENDED USE

This Immundiagnostik AG assay is an enzyme immunoassay intended for the quantitative determination of β_2 -microglobulin in serum, plasma and urine. For *in vitro* diagnostic use only.

2. INTRODUCTION

β_2 -microglobulin is a light chain protein (11.8 kD) of the HLA-class-I antigens and is found on the cell membrane of all nucleated cells. This protein is metabolised extensively in the kidney. The serum concentration is influenced by the rates of synthesis and metabolism and is usually stable in healthy persons. Changes in the serum concentrations are indicative of disorders in glomerular and tubular functions.

Indications

- Early detection of a renal transplant rejection
- Assessment of the glomerular filtration rate (GFR)

3. MATERIAL SUPPLIED

Cat. No.	Label	Kit components	Quantity
K 6210	PLATE	Microtiter plate, pre-coated	12 x 8 wells
K 0001.C.100	WASHBUF	Wash buffer concentrate, 10x	1 x 100 ml
K 6210	NACL	0.9 % NaCl-solution, ready-to-use	1 x 25 ml
K 6210	CONJ	Conjugate, ready-to-use (rabbit-anti- β_2 -microglobulin, peroxidase-labelled)	1 x 25 ml
K 6210	STD	Standards, lyophilised (0; 0.6; 1.2; 2.5; 5; 10 mg/l)	1 x 6 vials
K 6210	CTRL1	Control, lyophilised (see specification for range)	1 x 1 vial
K 6210	CTRL2	Control, lyophilised (see specification for range)	1 x 1 vial
K 6210	SAMPLEBUF	Sample dilution buffer, ready-to-use	1 x 100 ml
K 0002.15	SUB	Substrate (tetramethylbenzidine), ready-to-use	2 x 15 ml
K 0003.15	STOP	Stop solution, ready-to-use	1 x 15 ml

For reorders of single components, use the catalogue number followed by the label as product number.

4. MATERIAL REQUIRED BUT NOT SUPPLIED

- Ultrapure water*
- Calibrated precision pipettors and 10–1000 μ l single-use tips
- Foil to cover the microtiter plate
- Horizontal microtiter plate shaker
- Multi-channel pipets or repeater pipets
- Centrifuge
- Vortex
- Standard single-use laboratory glass or plastic vials, cups, etc.
- Microtiter plate reader (required filters see chapter 7)

* Immundiagnostik AG recommends the use of ultrapure water (water type 1; ISO 3696), which is free of undissolved and colloidal ions and organic molecules (free of particles > 0.2 μ m) with an electrical conductivity of 0.055 μ S/cm at 25 °C (\geq 18.2 M Ω cm).

5. STORAGE AND PREPARATION OF REAGENTS

- To run the assay more than once, ensure that reagents are stored at the conditions stated on the label. **Prepare only the appropriate amount necessary for each run.** The kit can be used up to 4 times within the expiry date stated on the label.
- **Preparation of the wash buffer:** The **wash buffer concentrate (WASHBUF)** has to be diluted with ultrapure water **1:10** before use (100 ml WASHBUF + 900 ml ultrapure water), mix well. Crystals could occur due to high salt concentration in the concentrate. Before dilution, the crystals have to be redissolved at room temperature or in a water bath at 37 °C. The **WASHBUF** is stable at **2–8 °C** until the expiry date stated on the label. **Wash buffer** (1:10 diluted WASHBUF) can be stored in a closed flask at **2–8 °C for 1 month.**
- The **lyophilised standards (STD)** and **controls (CTRL)** are stable at **2–8 °C** until the expiry date stated on the label. Before use, the STD and CTRL have to be reconstituted with **250 μ l of ultrapure water** and mixed by gentle inversion to ensure complete reconstitution. Allow the vial content to dissolve for 10 minutes and then mix thoroughly. **Standards and controls** (reconstituted STD and CTRL) **can be stored at 2–8 °C for 4 weeks. For long term storage up to 3 months they can be stored at -20 °C. Avoid repeated thawing and freezing.**
- All other test reagents are ready-to-use. Test reagents are stable until the expiry date (see label) when stored at **2–8 °C.**

6. STORAGE AND PREPARATION OF SAMPLES

Sample storage

Serum and Plasma

Samples can be stored for two weeks at 2-8 °C. For longer storage, samples should be frozen at -20 °C.

Urine

Urines should be adjusted to a pH of 6 to 8 with 1 N NaOH. Adjusted samples can be stored at 2-8 °C for 14 days. For longer storage, non-treated samples should be frozen at -20 °C.

Dilution of samples

Plasma, serum and urine samples must be diluted **1:50** before performing the assay, mix well. For example

10 μ l sample + **490 μ l** sample dilution buffer (SAMPLEBUF).

Samples with a **β_2 -microglobulin content higher than 10 mg/l** should be **further diluted 1:10 with sample dilution buffer**.

For analysis, pipet **10 μ l** of the dilution per well.

7. ASSAY PROCEDURE

Principle of the test

This ELISA is designed for the quantitative determination of β_2 -microglobulin.

In a first incubation step, β_2 -microglobulin is bound to an immobilised antibody. Then a peroxidase-labelled anti- β_2 -microglobulin-antibody is added and a sandwich of capture antibody – β_2 -microglobulin – peroxidase-conjugate is formed. Tetramethylbenzidine (TMB) is used as peroxidase substrate. Finally, an acidic stop solution is added to terminate the reaction. The colour changes from blue to yellow. The intensity of the colour is directly proportional to the concentration of β_2 -microglobulin. A dose response curve of the absorbance unit (optical density, OD at 450 nm) vs. concentration is generated, using the values obtained from the standards. β_2 -microglobulin, present in the patient samples, is determined directly from this curve.

Test procedure

Bring all **reagents and samples to room temperature** (15–30 °C) and mix well.

Mark the positions of standards/controls/samples on a protocol sheet.

Take as many microtiter strips as needed from the kit. Store unused strips together with the desiccant bag in the closed aluminium packaging at 2–8 °C. Strips are stable until expiry date stated on the label.

For automated ELISA processors, the given protocol may need to be adjusted according to the specific features of the respective automated platform. For further details please contact your supplier or Immundiagnostik AG.

We recommend to carry out the tests in duplicate.

1.	Before use , wash the wells 5 times with 250 µl wash buffer . After the final washing step, remove residual wash buffer by firmly tapping the plate on absorbent paper.
2.	Add 200 µl 0.9% NaCl-solution (NACL) into each well.
3.	Add each 10 µl standards/controls/diluted samples into the respective wells.
4.	Cover the strips and incubate for 1 hour at room temperature (15–30 °C) on a horizontal shaker* .
5.	Discard the content of each well and wash 5 times with 250 µl wash buffer . After the final washing step, remove residual wash buffer by firmly tapping the plate on absorbent paper.
6.	Add 200 µl conjugate (CONJ) into each well.
7.	Cover the strips and incubate for 15 min at room temperature (15–30 °C) on a horizontal shaker* .
8.	Discard the content of each well and wash 5 times with 250 µl wash buffer . After the final washing step, remove residual wash buffer by firmly tapping the plate on absorbent paper.
9.	Add 200 µl substrate (SUB) into each well.
10.	Incubate for 5–15 min** at room temperature (15–30 °C) in the dark .
11.	Add 50 µl stop solution (STOP) into each well and mix well.

12.	Determine absorption immediately with an ELISA reader at 450 nm against 620 nm (or 690 nm) as a reference. If no reference wavelength is available, read only at 450 nm. If the extinction of the highest standard exceeds the range of the photometer, absorption must be measured immediately at 405 nm against 620 nm as a reference.
-----	---

* We recommend shaking the strips at 550 rpm with an orbit of 2 mm.

** The intensity of the colour change is temperature sensitive. We recommend observing the colour change and stopping the reaction upon good differentiation.

8. RESULTS

The following algorithms can be used alternatively to calculate the results. We recommend using the 4 parameter algorithm.

1. 4 parameter algorithm

It is recommended to use a linear ordinate for the optical density and a logarithmic abscissa for the concentration. When using a logarithmic abscissa, the zero standard must be specified with a value less than 1 (e.g. 0.001).

2. Point-to-point calculation

We recommend a linear ordinate for the optical density and a linear abscissa for the concentration.

3. Spline algorithm

We recommend a linear ordinate for the optical density and a linear abscissa for the concentration.

The plausibility of the duplicate values should be examined before the automatic evaluation of the results. If this option is not available with the programme used, the duplicate values should be evaluated manually.

Serum, plasma and urine

Since the sample dilution is already considered in the calibration curve, the dilution-factor is 1.

In case **another dilution factor** has been used, multiply the obtained result by the dilution factor used.

9. LIMITATIONS

Samples with concentrations above the measurement range (see definition below) can be further diluted and re-assayed. Please consider this higher dilution when calculating the results.

Samples with concentrations lower than the measurement range (see definition below) cannot be clearly quantified.

The upper limit of the measurement range can be calculated as:

highest concentration of the standard curve \times sample dilution factor to be used

The lower limit of the measurement range can be calculated as:

Analytical sensitivity \times sample dilution factor to be used

Analytical sensitivity see chapter "Performance Characteristics".

10. QUALITY CONTROL

Immundiagnostik AG recommends the use of external controls for internal quality control, if possible.

Control samples should be analysed with each run. Results, generated from the analysis of control samples, should be evaluated for acceptability using appropriate statistical methods. The results for the patient samples may not be valid if within the same assay one or more values of the quality control sample are outside the acceptable limits.

Reference range

Plasma or serum:	< 2.5 mg/l
Urine:	< 0.4 mg/l

We recommend each laboratory to establish its own reference range.

11. PERFORMANCE CHARACTERISTICS

Spiking Recovery

A plasma and a urine sample each were spiked with 5 mg β_2 -microglobulin and measured using this assay. The recovery-rate for plasma was 98 % and for urine 93 %.

Dilution recovery

The linearity of this assay was estimated by dilution of a β_2 -microglobulin sample in sample dilution buffer. The linearity extends from 0.2 to 10 mg/l.

Precision and reproducibility

Intra-Assay (n = 12)

The reproducibility of two results in one measurement series was evaluated. A plasma and a urine sample were analysed 12 times by one person using the β_2 -microglobulin ELISA.

Sample	β_2 -microglobulin [mg/l]	CV [%]
urine	5.7	9
plasma	1.1	11

Inter-Assay (n = 12)

The reproducibility of two results at different days was evaluated. A plasma and a urine sample were analysed at different days by different persons using the β_2 -microglobulin ELISA.

Sample	β_2 -microglobulin [mg/l]	CV [%]
urine	5.9	15
plasma	1.0	12

Analytical Sensitivity

The detection limit was set as $B_0 + 2 \text{ SD}$ and estimated to be 0.1 mg/l.

12. PRECAUTIONS

- All reagents in the kit package are for *in vitro* diagnostic use only.
- Human materials used in kit components were tested and found to be negative for HIV, Hepatitis B and Hepatitis C. However, for safety reasons, all kit components should be treated as potentially infectious.
- Kit reagents contain sodium azide or ProClin as bactericides. Sodium azide and ProClin are toxic. Substrates for the enzymatic colour reactions are toxic and carcinogenic. Avoid contact with skin or mucous membranes.
- The stop solution consists of diluted sulphuric acid, a strong acid. Although diluted, it still must be handled with care. It can cause burns and should be handled with gloves, eye protection, and appropriate protective clothing. Any

spill should be wiped up immediately with copious quantities of water. Do not breath vapour and avoid inhalation.












13. TECHNICAL HINTS

- Do not interchange different lot numbers of any kit component within the same assay. Furthermore we recommend not assembling wells of different microtiter plates for analysis, even if they are of the same batch.
- Control samples should be analysed with each run.
- Reagents should not be used beyond the expiration date stated on the kit label.
- Substrate solution should remain colourless until use.
- To ensure accurate results, proper adhesion of plate sealers during incubation steps is necessary.
- Avoid foaming when mixing reagents.
- Do not mix plugs and caps from different reagents.
- The assay should always be performed according to the enclosed manual.

14. GENERAL NOTES ON THE TEST AND TEST PROCEDURE

- This assay was produced and distributed according to the IVD guidelines of 98/79/EC.
- The guidelines for medical laboratories should be followed.
- Incubation time, incubation temperature and pipetting volumes of the components are defined by the producer. Any variation of the test procedure, which is not coordinated with the producer, may influence the results of the test. Immundiagnostik AG can therefore not be held responsible for any damage resulting from incorrect use.
- Warranty claims and complaints regarding deficiencies must be logged within 14 days after receipt of the product. The product should be send to Immundiagnostik AG along with a written complaint.

Used symbols:

	Temperature limitation		Catalogue Number
	In Vitro Diagnostic Medical Device		To be used with
	Manufacturer		Contains sufficient for <n> tests
	Lot number		Use by
	Attention		Consult instructions for use
	Consult specification data sheet		