



Distribuito in ITALIA da
Li StarFish S.r.l.
Via Cavour, 35
20063 Cernusco S/N (MI)
telefono 02-92150794
info@listarfish.it
www.listarfish.it


Manual

IDK[®] Hb/Hp complex ELISA

***For the in vitro determination of hemoglobin/haptoglobin
in stool***

Valid from 2019-01-31

REF **K 7817D**



IVD **CE**



Immundiagnostik AG, Stubenwald-Allee 8a, 64625 Bensheim, Germany

Tel.: +49 6251 70190-0

Fax: + 49 6251 70190-363

e.mail: info@immundiagnostik.com www.immundiagnostik.com

Table of Contents

1. INTENDED USE	17
2. INTRODUCTION	17
3. MATERIAL SUPPLIED	17
4. MATERIAL REQUIRED BUT NOT SUPPLIED	18
5. PREPARATION AND STORAGE OF REAGENTS	18
6. STORAGE AND PREPARATION OF SAMPLES	19
<i>Extraction of the stool samples</i>	19
<i>Sample stability and storage</i>	20
7. ASSAY PROCEDURE	20
<i>Principle of the test</i>	20
<i>Test procedure</i>	21
8. RESULTS	22
9. LIMITATIONS	23
10. QUALITY CONTROL	23
<i>Reference range</i>	23
11. PERFORMANCE CHARACTERISTICS	24
<i>Accuracy – Precision</i>	24
<i>Analytical specificity</i>	24
<i>Linearity</i>	25
<i>Accuracy – Trueness</i>	25
<i>Analytical sensitivity</i>	26
12. PRECAUTIONS	26
13. TECHNICAL HINTS	26
14. GENERAL NOTES ON THE TEST AND TEST PROCEDURE	27
15. REFERENCES	27
<i>General literature</i>	27
<i>Literature using the IDK® Hb/Hp complex ELISA</i>	27

1. INTENDED USE

This Immundiagnostik AG assay is an enzyme immunoassay intended for the quantitative determination of human **hemoglobin/haptoglobin complex** in stool. For *in vitro* diagnostic use only.

2. INTRODUCTION

In contrast to commercially available rapid tests, this hemoglobin/haptoglobin complex ELISA does not require previous adherence to a diet (no raw meat etc.) and recognises human hemoglobin/haptoglobin complex in 100-fold lower concentrations. This avoids false-negative results. Because of the antibodies used, false-positive results are almost excluded. Recent data show that the hemoglobin/haptoglobin complex determination increases the clinical specificity and sensitivity.

Indications

- Occult blood in stool
- Crohn's disease; Ulcerative Colitis
- Suspicion of colon carcinoma
- Polyps in the colorectum

3. MATERIAL SUPPLIED

Cat. No.	Label	Kit components	Quantity
K 7817D	PLATE	Microtiter plate, pre-coated	12 x 8 wells
K 0001.C.100	WASHBUF	Wash buffer concentrate, 10x	2 x 100 ml
K 6999.C.100	IDK Extract®	Extraction buffer concentrate <i>IDK Extract®</i> , 2.5 x	1 x 100 ml
K 7817D	SAMPLEBUF	Sample dilution buffer, ready-to-use	2 x 15 ml
K 7817D	CONJ	Conjugate, ready-to-use, peroxidase-labelled (mouse-anti humanHb)	1 x 15 ml
K 7817D	STD	Standards, lyophilised (50; 10; 3,3; 0,67; 0 µg/g)	4 x 5 vials
K 7817D	CTRL1	Control, lyophilised (see specification for range)	4 x 1 vials
K 7817D	CTRL2	Control, lyophilised (see specification for range)	4 x 1 vials

Cat. No.	Label	Kit components	Quantity
K 0002.15	SUB	Substrate (Tetramethylbenzidine), ready-to-use	1 x 15 ml
K 0003.15	STOP	Stop solution, ready-to-use	1 x 15 ml

For reorders of single components, use the catalogue number followed by the label as product number.

4. MATERIAL REQUIRED BUT NOT SUPPLIED

- Ultrapure water*
- Stool sample application system such as Cat. No.: K 6998SAS
- Calibrated precision pipettors and 10–1000 µl single-use tips
- Foil to cover the microtiter plate
- Horizontal microtiter plate shaker
- Multi-channel pipets or repeater pipets
- Centrifuge, 3000 g
- Vortex
- Standard single-use laboratory glass or plastic vials, cups, etc.
- Microtiter plate reader (required filters see chapter 7)

* Immundiagnostik AG recommends the use of ultrapure water (water type 1; ISO 3696), which is free of undissolved and colloidal ions and organic molecules (free of particles > 0.2 µm) with an electrical conductivity of 0.055 µS/cm at 25 °C (≥ 18.2 MΩ cm).

5. PREPARATION AND STORAGE OF REAGENTS

- To run the assay more than once, ensure that reagents are stored at the conditions stated on the label. **Prepare only the appropriate amount necessary for each run.** The kit can be used up to 4 times within the expiry date stated on the label.
- **Preparation of the wash buffer:** The **wash buffer concentrate (WASHBUF)** has to be diluted with ultrapure water **1:10** before use (100 ml WASHBUF + 900 ml ultrapure water), mix well. Crystals could occur due to high salt concentration in the concentrate. Before dilution, the crystals have to be redissolved at room temperature or in a water bath at 37 °C. The **WASHBUF** is stable at **2–8 °C** until the expiry date stated on the label. **Wash buffer** (1:10 diluted WASHBUF) can be stored in a closed flask at **2–8 °C for 1 month**.
- **Preparation of the extraction buffer:** The **extraction buffer concentrate IDK Extract®** has to be diluted with ultrapure water **1:2.5** before use (100 ml **IDK Extract®** + 150 ml ultrapure water), mix well. Crystals could occur due to

high salt concentration in the stock solutions. Before dilution, the crystals have to be redissolved at 37°C in a water bath. The **IDK Extract®** is stable at **2–8 °C** until the expiry date stated on the label. Extraction buffer (1:2.5 diluted **IDK Extract®**) can be stored in a closed flask at **2–8 °C for four months**.

- The **lyophilised standards (STD)** and **controls (CTRL)** are stable at **2–8 °C** until the expiry date stated on the label. Before use, the STD and CTRL have to be reconstituted with **500 µl of ultrapure water** and mixed by gentle inversion to ensure complete reconstitution. Allow the vial content to dissolve for 10 minutes and then mix thoroughly. **Standards and controls** (reconstituted STD and CTRL) **are not stable and cannot be stored**.
- All other test reagents are ready-to-use. Test reagents are stable until the expiry date (see label) when stored at **2–8 °C**.

6. STORAGE AND PREPARATION OF SAMPLES

Extraction of the stool samples

Extraction buffer (1:2.5 diluted **IDK Extract®**) is used as a sample extraction buffer. We recommend the following sample preparation:

Stool Sample Application System (SAS) (Cat. No.: K 6998SAS)

Stool sample tube – Instructions for use

Please note that the dilution factor of the final stool suspension depends on the amount of stool sample used and the volume of the buffer.

SAS with 1.5 ml sample extraction buffer:

Applied amount of stool:	15 mg
Buffer Volume:	1.5 ml
Dilution Factor:	1:100

Please follow the instructions for the preparation of stool samples using the SAS as follows:

- a) The raw stool sample has to be thawed. For particularly heterogeneous samples we recommend a mechanical homogenisation using an applicator, inoculation loop or similar device.
- b) Fill the **empty stool sample tube** with **1.5 ml sample extraction buffer** (1:2.5 diluted **IDK Extract®**) before using it with the sample. **Important:** Allow the sample extraction buffer to reach room temperature.

- c) Unscrew the tube (yellow part of cap) to open. Insert the yellow dipstick into the sample. The lower part of the dipstick has notches which need to be covered completely with stool after inserting it into the sample. Place dipstick back into the tube. When putting the stick back into the tube, excess material will be stripped off, leaving 15 mg of sample to be diluted. Screw tightly to close the tube.
- d) Vortex the tube well until no stool sample remains in the notches. **Important:** Please make sure that you have a maximally homogenous suspension after shaking. Especially with more solid samples, soaking the sample in the tube with sample extraction buffer for ~ 10 minutes improves the result.
- e) Allow sample to stand for ~10 minutes until sediment has settled. Floating material like shells of grains can be neglected.
- f) Carefully unscrew the complete cap of the tube including the blue ring plus the dipstick. Discard cap and dipstick. Make sure that the sediment will not be dispersed again.

Dilution I: 1:100

The sample suspension is now ready for use.

The sample can also be used in a pipetting automat. Place the sample in the sample rack according to instrument instructions.

Sample stability and storage

Due to the degradation of hemoglobin/haptoglobin at room temperature, which can amount to 50% per day, **raw stool** samples should be shipped frozen. If shipment either at -20 °C or cooled is not possible, the samples can be mailed overnight, but this will reduce the sensitivity. Raw stool can be stored at -20 °C for 1 month.

Stool extract is stable at room temperature (15–30 °C), at 2–8 °C as well as at -20 °C for seven days. Avoid more than three freeze-thaw cycles.

7. ASSAY PROCEDURE

Principle of the test

This ELISA is designed for the quantitative determination of hemoglobin/haptoglobin complex in stool. The complex in the sample is bound to anti-haptoglobin antibodies (in excess), which are immobilised on the surface of the microtiter wells. To remove all unbound substances, a washing step is carried out. In a second incubation step an anti-hemoglobin peroxidase labelled antibody is added. After another

washing step, to remove all unbound substances, the solid phase is incubated with the substrate, tetramethylbenzidine (TMB). An acidic solution is then added to stop the reaction. The colour converts from blue to yellow. The intensity of the yellow colour is directly proportional to the concentration of hemoglobin/haptoglobin complex in the sample. A dose response curve of the absorbance unit (optical density, OD) vs. concentration is generated, using the values obtained from standard. Hemoglobin/haptoglobin complex, present in the patient samples, is determined directly from this curve.

Test procedure

Bring all **reagents and samples to room temperature** (15–30°C) and mix well.

Mark the positions of standards/controls/samples on a protocol sheet.

Take as many microtiter strips as needed from kit. Store unused strips together with the desiccant bag in the closed aluminium packaging at 2–8°C. Strips are stable until expiry date stated on the label.

For automated ELISA processors, the given protocol may need to be adjusted according to the specific features of the respective automated platform. For further details please contact your supplier or Immundiagnostik AG.

We recommend to carry out the tests in duplicate.

1.	Before use , wash the wells 5 times with 250 µl wash buffer . After the final washing step, remove residual wash buffer by firmly tapping the plate on absorbent paper.
2.	Add 50 µl of sample dilution buffer (SAMPLEBUF) into each well.
3.	Add each 50 µl standards/controls/diluted samples into the respective wells.
4.	Cover the strips and incubate for 1 hour at room temperature (15–30°C).
5.	Discard the content of each well and wash 5 times with 250 µl wash buffer . After the final washing step, remove residual wash buffer by firmly tapping the plate on absorbent paper.
6.	Add 100 µl conjugate (CONJ) into each well.
7.	Cover the strips and incubate for 1 hour at room temperature (15–30°C).
8.	Discard the content of each well and wash 5 times with 250 µl wash buffer . After the final washing step, remove residual wash buffer by firmly tapping the plate on absorbent paper.

9.	Add 100 µl substrate (SUB) into each well.
10.	Incubate for 10–20 min* at room temperature (15–30 °C) in the dark .
11.	Add 100 µl stop solution (STOP) into each well and mix well.
12.	Determine absorption immediately with an ELISA reader at 450 nm against 620 nm (or 690 nm) as a reference. If no reference wavelength is available, read only at 450 nm. If the extinction of the highest standard exceeds the range of the photometer, absorption must be measured immediately at 405 nm against 620 nm as a reference.

* The intensity of the colour change is temperature sensitive. We recommend observing the colour change and stopping the reaction upon good differentiation.

8. RESULTS

The following algorithms can be used alternatively to calculate the results. We recommend using the „4 parameter algorithm“.

1. 4 parameter algorithm

It is recommended to use a linear ordinate for the optical density and a logarithmic abscissa for the concentration. When using a logarithmic abscissa, the zero standard must be specified with a value less than 1 (e.g. 0.001).

2. Point-to-point calculation

We recommend a linear ordinate for the optical density and a linear abscissa for the concentration.

3. Spline algorithm

We recommend a linear ordinate for the optical density and a linear abscissa for the concentration.

The plausibility of the pairs of values should be examined before the automatic evaluation of the results. If this option is not available with the used program, a control of the paired values should be done manually.

Stool

Since the sample dilution is already considered in the calibration curve, the dilution factor is 1.

In case **another dilution factor** has been used, multiply the obtained result by the dilution factor used.

9. LIMITATIONS

Samples with concentrations above the measurement range (see definition below) can be further diluted and re-assayed. Please consider this higher dilution when calculating the results.

Samples with concentrations lower than the measurement range (see definition below) cannot be clearly quantified.

The upper limit of the measurement range can be calculated as:

highest concentration of the standard curve × sample dilution factor to be used

The lower limit of the measurement range can be calculated as:

LoB × sample dilution factor to be used

LoB see chapter "Performance Characteristics".

10. QUALITY CONTROL

Immundiagnostik AG recommends the use of external controls for internal quality control, if possible.

Control samples should be analysed with each run. Results, generated from the analysis of control samples, should be evaluated for acceptability using appropriate statistical methods. The results for the patient samples may not be valid if within the same assay one or more values of the quality control sample are outside the acceptable limits.

Reference range

Hemoglobin/haptoglobin (stool)*: **< 2 µg/ml ≙ < 2 µg/g**
(1 g stool ≙ 1 ml)

* H.G. Bischoff et al., 1990. Fecal Occult Blood: A Sensitive and Specific Method for Detection by Immunological Determination of Human Albumin and Hemoglobin. *Ärztl. Lab.*, **36**, pp. 101–112.

We recommend each laboratory to establish its own reference range.

11. PERFORMANCE CHARACTERISTICS

Accuracy – Precision

Repeatability (Intra-Assay); n = 42

The repeatability was assessed with 2 stool samples under **constant** parameters (same operator, instrument, day and kit lot).

Sample	Mean value [$\mu\text{g/g}$]	CV [%]
1	8.31	3.9
2	3.88	2.5

Reproducibility (Inter-Assay); n = 30

The reproducibility was assessed with 2 control samples under **varying** parameters (different operators, instruments, days and kit lots).

Sample	Mean value [$\mu\text{g/g}$]	CV [%]
1	1.16	4.5
2	4.20	2.7

Analytical specificity

The specificity of the antibody was tested by measuring the cross-reactivity against 5 compounds with structural similarity to hb-hp-complex. There was no cross-reactivity observed.

Substance tested	Concentration added	Concentration obtained [$\mu\text{g/g}$]	Conclusion
Pankreatic Amylase	28 000 mU/l	0.000	< LoB
MPO	100 ng/ml	0.016	< LoB
Lysozyme	500 ng/ml	0.006	< LoB
Chymotrypsin	1 000 ng/ml	0.011	< LoB
slgA	600 ng/ml	0.008	< LoB

Linearity

The linearity states the ability of a method to provide results proportional to the concentration of analyte in the test sample within a given range. This was assessed according to CLSI guideline EP06-A by serial dilution of 2 different stool samples.

For hb-hp-complex in stool, the method has been demonstrated to be linear from 1.09 to 24.86 µg/g, showing a non-linear behaviour of less than ±20 % in this interval.

Sample	Dilution	Expected [µg/g]	Obtained [µg/g]	Recovery [%]
A	undiluted	24.86	24.86	100.00
	1:2	12.43	12.91	103.87
	1:4	6.21	7.10	114.25
	1:8	3.11	3.64	117.05
	1:16	1.55	1.82	117.09
B	undiluted	8.69	8.69	100.00
	1:2	4.34	4.28	101.40
	1:4	2.17	2.45	88.73
	1:8	1.09	1.27	85.58

Accuracy – Trueness

The trueness states the closeness of the agreement between the result of a measurement and the true value of the measurand. Therefore, hb-hp-complex-spikes with known concentrations were added to 2 different stool samples.

Sample	Spike [µg/g]	Obtained [µg/g]	Expected [µg/g]	Recovery [%]
0.414	9.55	9.93	10.26	103.36
	5.30	5.70	6.62	116.17
	2.70	3.11	3.72	119.63
	1.20	1.61	1.87	115.86
0.922	9.55	10.40	8.40	80.75
	5.30	6.19	5.39	87.10
	2.70	3.61	3.74	103.80
	1.20	2.12	2.03	95.90

Analytical sensitivity

The following values have been estimated based on the concentrations of the standard without considering possibly used sample dilution factors

Limit of blank, LoB	0.033 µg/g
Limit of detection, LoD	0.158 µg/g
Limit of quantitation, LoQ	0.248 µg/g

The evaluation was performed according to the CLSI guideline EP-17-A2. The specified accuracy goal for the LoQ was 20 % CV.

12. PRECAUTIONS

- All reagents in the kit package are for *in vitro* diagnostic use only.
- Human materials used in kit components were tested and found to be negative for HIV, Hepatitis B and Hepatitis C. However, for safety reasons, all kit components should be treated as potentially infectious.
- Kit reagents contain sodium azide or Proclin as bactericides. Sodium azide and Proclin are toxic. Substrates for the enzymatic colour reactions are toxic and carcinogenic. Avoid contact with skin or mucous membranes.
- The stop solution consists of diluted sulphuric acid, a strong acid. Although diluted, it still must be handled with care. It can cause burns and should be handled with gloves, eye protection, and appropriate protective clothing. Any spill should be wiped up immediately with copious quantities of water. Do not breath vapour and avoid inhalation.

13. TECHNICAL HINTS

- Do not interchange different lot numbers of any kit component within the same assay. Furthermore we recommend not assembling wells of different microtiter plates for analysis, even if they are of the same batch.
- Control samples should be analysed with each run.
- Reagents should not be used beyond the expiration date stated on kit label.
- Substrate solution should remain colourless until use.
- To ensure accurate results, proper adhesion of plate sealers during incubation steps is necessary.
- Avoid foaming when mixing reagents.

- Do not mix plugs and caps from different reagents.
- The assay should always be performed according to the enclosed manual.

14. GENERAL NOTES ON THE TEST AND TEST PROCEDURE

- This assay was produced and distributed according to the IVD guidelines of 98/79/EC.
- The guidelines for medical laboratories should be followed.
- *IDK®* and *IDK Extract®* are trademarks of Immundiagnostik AG.
- Incubation time, incubation temperature and pipetting volumes of the components are defined by the producer. Any variation of the test procedure, which is not coordinated with the producer, may influence the results of the test. Immundiagnostik AG can therefore not be held responsible for any damage resulting from incorrect use.
- Warranty claims and complaints regarding deficiencies must be logged within 14 days after receipt of the product. The product should be sent to Immundiagnostik AG along with a written complaint.

15. REFERENCES












General literature

1. Lüthgens, K. et al., 1998. Hemoglobin-Haptoglobin-Complex: A Highly Sensitive Assay for the Detection of Fecal Occult Blood. *Clinical laboratory*, **44**, pp.543–551.
2. Thomas, L., 1998. Labor und Diagnose. Indikation und Bewertung von Laborbefunden für die medizinische Diagnostik **5th ed.**, Frankfurt/Main: *TH-Books Verlagsgesellschaft*.
3. John, M. et al., 1994. Nachweis von Albumin im Stuhl zur Erkennung okkultur Blutungen : Vergleich zweier immunologischer Tests . Radiale Immundiffusion vs BM-Test Colon Albumin. *Klinisches Labor*, **40**, pp.77–81.

Literature using the IDK® Hb/Hp complex ELISA

4. Chalkias, A. et al., 2011. Patients with colorectal cancer are characterized by increased concentration of fecal hb-hp complex, myeloperoxidase, and secretory IgA. *American journal of clinical oncology*, **34**(6), pp.561–6.

Used symbols:

	Temperature limitation		Catalogue Number
	In Vitro Diagnostic Medical Device		To be used with
	Manufacturer		Contains sufficient for <n> tests
	Lot number		Use by
	Attention		Consult instructions for use
	Consult specification data sheet		