



Distribuito in ITALIA da  
**Li StarFish S.r.l.**  
Via Cavour, 35  
20063 Cernusco S/N (MI)  
telefono 02-92150794  
info@listarfish.it  
www.listarfish.it

**Manual**


## **ID-Vit<sup>®</sup> vitamin B<sub>6</sub>**

***Microbiological test kit for the determination of vitamin B<sub>6</sub> in serum using a *Saccharomyces cerevisiae* coated microtitre plate***

***For use in human and veterinary medicine and in research***

Valid from 2019-05-22

**REF** KIF006



**IVD** **CE**



**Immundiagnostik AG**, Stubenwald-Allee 8a, 64625 Bensheim, Germany  
Tel.: +49 6251 70190-0 Fax: +49 6251 70190-363  
e.mail: info@immundiagnostik.com www.immundiagnostik.com

# Table of Contents

<b>1. INTENDED USE</b>	<b>17</b>
<b>2. INTRODUCTION</b>	<b>17</b>
<b>3. PRINCIPLE OF THE TEST</b>	<b>18</b>
<b>4. MATERIAL SUPPLIED</b>	<b>18</b>
<b>5. MATERIAL REQUIRED BUT NOT SUPPLIED</b>	<b>18</b>
<b>6. PRECAUTIONS</b>	<b>19</b>
<b>7. STORAGE AND PREPARATION OF REAGENTS</b>	<b>19</b>
7.1 Water	19
7.2 Preparation of the enzyme solution	20
7.3 Preparation of the controls	20
7.4 Preparation of the standard curve	20
7.5 Preparation of the sterile assay medium	21
7.6 Microtiter plate [PLATE]	21
<b>8. SAMPLE STORAGE AND PREPARATION</b>	<b>21</b>
8.1 Sample pretreatment	22
8.2 Sample dilution	22
<b>9. ASSAY PROCEDURE</b>	<b>22</b>
9.1 Test preparations	22
9.2 Test procedure	22
9.3 Measurement	23
<b>10. EVALUATION OF RESULTS</b>	<b>23</b>
10.1 Calculation	23
10.2 Quality control	24
<b>11. LIMITATIONS</b>	<b>24</b>
<b>12. PERFORMANCE CHARACTERISTICS</b>	<b>24</b>
12.1 Precision and reproducibility	24
12.2 Recovery	25
<b>13. REFERENCES</b>	<b>26</b>
<b>14. GENERAL NOTES ON THE TEST AND TEST PROCEDURE</b>	<b>26</b>

## 1. INTENDED USE

ID-Vit® Vitamin B<sub>6</sub> is a microtiter plate test kit based on a microbiological method which measures the vitamin B<sub>6</sub> content in serum. The test kit contains the standard and all reagents required to perform the test. It is sufficient for 96 determinations including standard curves. An ELISA reader is required for the evaluation of the results. For use in human and veterinary medicine and in research. For *in vitro* diagnostic use only.

## 2. INTRODUCTION

The vitamin B<sub>6</sub> group comprises three natural forms: pyridoxine, pyridoxamine and pyridoxal, which, during metabolism, are converted to the enzymatically active form pyridoxal-5-phosphate (P5P). Pyridoxal-5-phosphate (here described as „vitamin B<sub>6</sub>“) is a cofactor in more than a hundred enzyme reactions. One of its functions is to perform transamination, a key step in both breaking down and producing amino acids in the body. Vitamin B<sub>6</sub> is also necessary for the synthesis of neurotransmitters, as well as hemoglobin in red blood cells. And vitamin B<sub>6</sub> plays a central role in fat metabolism.

### Vitamin B<sub>6</sub> deficiency symptoms

- Disturbance of protein biosynthesis
- Muscular weakness, loss of muscle control
- Skin disorders (dermatitis, pigment abnormality)
- Nervous disorders (irritability, depression, palsy)
- Insomnia

Vitamin B<sub>6</sub> deficiency can be considered as a risk factor for myocardial infarction, peripheral vascular diseases and atherosclerosis, especially in connection with the regulation of homocysteine metabolism.

### Indications

Vitamin B<sub>6</sub> deficiency can result from:

- Chronic inflammatory bowel diseases (colitis ulcerosa, Crohn´s disease, gluten sensitivity)
- Dialysis
- Alcohol abuse
- Homocysteinuria
- Pregnancy and lactation

### 3. PRINCIPLE OF THE TEST

The serum samples are enzymatically pre-treated and then transferred into the wells of a microtiter plate coated with *Saccharomyces cerevisiae*. The addition of vitamin B<sub>6</sub> in either standards or samples gives a vitamin B<sub>6</sub>-dependent growth response until vitamin B<sub>6</sub> is consumed. After incubation at **30°C** for **48 h**, the growth of *Saccharomyces cerevisiae* is measured turbidimetrically at 610–630 nm (alternatively at 540–550 nm) in an ELISA reader and compared to a standard curve generated from the dilution series. The amount of vitamin B<sub>6</sub> is directly proportional to the turbidity.

### 4. MATERIAL SUPPLIED

Cat. No.	Label	Kit components	Quantity
KIF006	PLATE	<i>Saccharomyces cerevisiae</i> -precoated microtiter plate	1 x
KIF006	SOL	Sample preparation buffer	4 x 5 ml
KIF006	ENZ	Enzyme, lyoph.	4 x
KIF006	DIL	Water	4 x 30 ml
KIF006	ASYMED	Vitamin B <sub>6</sub> assay medium	4 x
KIF006	STD	Vitamin B <sub>6</sub> standard, lyoph.	4 x
KIF006	FOL	Adhesive cover foil	4 x
KIF006	FRA	Replacement holder for microtiter strips	1 x
KIF006	CTRL1	Vitamin B <sub>6</sub> control 1, lyoph.	4 x
KIF006	CTRL2	Vitamin B <sub>6</sub> control 2, lyoph.	4 x

### 5. MATERIAL REQUIRED BUT NOT SUPPLIED

- Incubator with a dark incubation chamber, 30 °C
- Water bath (90 °C–100 °C)
- Water bath or thermoblock with variable temperatur (37 °C, 95 °C)
- ELISA reader 610–630 nm (540–550 nm)
- Calibrated precision pipettors and sterile 20–1000 µl tips
- 5 ml and 10 ml pipets
- 1.5–2 ml reaction vials, sterile
- 0.2 µm sterile polyethersulfone (PES) filter with a sterile disposable syringe
- 15 ml centrifuge tubes, sterile (e.g. Falcon tubes)
- Biocentrifuge (10 000 g)

## 6. PRECAUTIONS

- As the test is based on a microbiological method, the general guidelines for sterile work should be observed as far as possible (preferably work in a sterile bench / PCR hood, use of sterile instruments or equipment).
- GLP (Good Laboratory Practice) guidelines have to be observed.
- Water quality is extremely important for the test. Only the water delivered with the test kit [DIL] should be used.
- For sterile filtration, only a sterile polyethersulfone filter must be used.
- It is essential to run a standard curve for each separate assay.
- Controls should be measured with each assay.
- We recommend measurements in duplicate.
- If a higher dilution results in a higher value measured, inhibitors like antibiotics might be present.
- Reagents should not be used beyond the expiration date shown on the label.
- Wear gloves during the test.
- Used microtiter stripes [PLATE] and materials that have been in contact with patient samples should be handled and disposed as potentially infectious.

## 7. STORAGE AND PREPARATION OF REAGENTS

- Store test kit and reagents at 2–8°C.
- Prepare reagents freshly and use them immediately after preparation. Discard remaining unused reagents and waste in accordance with country, federal, state, and local regulations.
- To run the assay more than once, ensure that reagents are stored at the conditions stated on the label. Prepare only the appropriate amount necessary for each run. The kit can be used up to 4 times within the expiry date stated on the label.

### 7.1 Water

- Water [DIL] (for medium [ASYMED], standard [STD] and controls [CTRL1, CTRL2])
- Push the lid up and pull it back to the rim of the glass, then twist the whole cap off.

## 7.2 Preparation of the enzyme solution

- Add 4 ml sample preparation buffer [SOL] to a vial of lyophilised enzyme [ENZ], then homogenise using a vortex.
- Enzyme solution cannot be stored.

## 7.3 Preparation of the controls

- The lyophilised controls [CTRL1, CTRL2] have to be resuspended with each 750 µl water [DIL] from the test kit, then homogenise using a vortex.
- After reconstitution, the controls are treated like samples.
- The concentration of the controls changes from lot to lot and is stated in the product specification.

## 7.4 Preparation of the standard curve

- For the preparation of the standard curve, standard concentrate is needed. To prepare standard concentrate, resuspend the lyophilised standard [STD] with x ml (x = please see the enclosed quality control protocol for the volume needed) water [DIL] supplied with the test kit, then homogenise using a vortex.
- Prepare a standard curve in 6 sterile reaction tubes (1.5–2 ml volume) from standard concentrate and water [DIL] following the scheme depicted in the table below:

Vitamin B <sub>6</sub> [µg/l]	Water [DIL] [µl]	+	Standard concentrate [µl]	=	Total volume [µl]
Blank: 0	940	+	0	=	940
Standard 1: 0.36	940	+	60	=	1000
Standard 2: 1.2	400	+	100	=	500
Standard 3: 1.8	350	+	150	=	500
Standard 4: 2.4	300	+	200	=	500
Standard 5: 3.6	200	+	300	=	500

### 7.5 Preparation of the sterile assay medium

- Fresh sterile assay medium has to be prepared each time before performing a test.
- Remove lyophilised assay medium from the desiccant bag in the assay medium bottle by taking the bag with a forceps and shaking it whilst still inside the bottle. Then remove the clean desiccant bag and discard it.
- Add 10 ml water [DIL] to the assay medium bottle [ASYMED], close the bottle firmly and shake it. This amount is sufficient for 6 microtiter stripes.
- Heat the medium bottle in a water bath at 90–100 °C for 5 min, shake well at least 2 times during this incubation time. Take care that the medium bottle is always firmly closed.
- Quickly cool the medium bottle to < 30 °C (at 2–8 °C for 10 min).
- Filter the medium using a disposable syringe (10 ml) and the 0.2 µm PES filter into a sterile centrifuge tube (15 ml, e. g. Falcon).
- After this preparation, the sterile assay medium can be used in the test.

### 7.6 Microtiter plate [PLATE]

- Store the microtiter plate [PLATE] in the aluminium packaging containing the desiccant bag at 2–8 °C.
- The microtiter plate [PLATE] has to be protected from humidity and contamination.
- Take care that the aluminium packaging is not damaged.
- Carefully close the aluminium packaging after opening.
- Take only the microtiter stripes needed directly before usage to avoid contamination

## 8. SAMPLE STORAGE AND PREPARATION

- Use serum for analysis.
- Samples are stable at 2–8 °C for one day in the dark. For longer storage, samples should be frozen and kept at -20 °C.

- Hemolytic samples may give erroneous results and should not be used for analysis. Lipemic samples should be centrifuged at 13 000 *g* before assaying to obtain fat free serum as far as possible.
- Samples should be centrifuged (at least 5 min at 10 000 *g*) prior to measurement. Use the resulting supernatant in the test.

### 8.1 Sample pretreatment

Add 300 µl serum/control to 300 µl of prepared enzyme solution (ratio 1:2), mix, and incubate at 37 °C for 30 min in the dark. Then heat to 95 °C for 30 min, cool quickly (at 2–8 °C for 10 min) and centrifuge for 10 min at 10 000 *g*.

### 8.2 Sample dilution

Take 100 µl from the supernatant of the prepared sample/control, add 400 µl water [DIL] and mix. The sample treatment and dilution result in a total dilution of 1:10 (= sample dilution factor).

## 9. ASSAY PROCEDURE

### 9.1 Test preparations

Take as many microtiter strips as needed from kit. Return unused strips and any unused test kit component to the original packaging, and put in the refrigerator. Bring all necessary reagents to room temperature.

### 9.2 Test procedure

- Take as many microtiter strips as needed from the kit and put them in the second microtiter strip holder [FRA].
- Put 150 µl sterile assay medium into the cavities.
- Add each 150 µl of the prepared standard curve, samples and controls into the respective cavities. Pre-rinse each pipet tip with standard, control or sample solution, respectively.
- Carefully seal the plate with adhesive cover foil [FOL]. Important: the cavities must be made airtight by pressing the foil down with the hand!
- Keep at **30 °C** for 48 h in an incubator.



### 9.3 Measurement

- Press the adhesive cover foil [FOL] firmly down again with the hand.
- Upturn the microtiter plate [PLATE], put it onto a tabletop and shake the microbes well.
- Turn the microtiter plate [PLATE] over again and carefully remove the adhesive cover foil [FOL]. During this, fix the strips in the frame with your hand because the foil is highly adhesive.
- Remove air bubbles in the cavities using a pipet tip or a needle.
- Read turbidity in an ELISA reader at E 610–630 nm (alternatively at E 540–550 nm).

#### Please note

- After 48 h incubation time, the microtiter plate [PLATE] may be stored for a maximum of 48 h in the refrigerator before measuring the turbidity.
- To prevent time-loss through public holidays or weekends, the microtiter plate [PLATE] may also be evaluated after 60 h incubation.

## 10. EVALUATION OF RESULTS

We recommend to use the 4 parameter algorithm to calculate the results. The sample dilution factor has to be considered for data evaluation.

The blank should have an optical density < standard 1. It serves as optical control to exclude contaminations and is not included in the calculation of results.

### 10.1 Calculation

Vitamin B<sub>6</sub> in µg/l = value from the standard curve × sample dilution factor (10)

#### Reference value for human serum

Based on studies of serum samples of apparently healthy persons (n = 74), the following values were estimated.

Vitamin B<sub>6</sub>: 4.8–17.7 µg/l

#### Please note

A concentration range of 3.6–36.1 µg/l vitamin B<sub>6</sub> is covered at a sample dilution of 1:10.

We recommend each laboratory to develop its own normal range as normal ranges strongly depend on the choice of the patient collective. The values mentioned above are only for orientation and can deviate from other published data.

## 10.2 Quality control

The extinction of the highest standard has to be > 0.6.

Results, generated from the analysis of control samples, should be evaluated for acceptability. The results for the samples may not be valid if within the same assay one or more values of the quality control sample or the highest standard are outside the acceptable limits.

## 11. LIMITATIONS

Whole blood cannot be used in the assay.

## 12. PERFORMANCE CHARACTERISTICS

The following performance characteristics have been collected using human serum samples.

### 12.1 Precision and reproducibility

#### Intraassay (n = 16)

	Vitamin B <sub>6</sub> [µg/l]	CV [%]
Sample	9.87	9.17

#### Interassay (n = 5)

	Vitamin B <sub>6</sub> [µg/l]	CV [%]
Sample	10.98	6.78

## 12.2 Recovery

Samples from 3 patients were spiked with vitamin B<sub>6</sub> and analysed. The mean values are shown below.

Sample (n=5)	Mean value original sample [µg/l]	Spike [µg/l]	Vitamin B <sub>6</sub> expected [µg/l]	Vitamin B <sub>6</sub> measured [µg/l]	Recovery Rate [%]
A	11.99	10	21.99	26.00	140
		20	31.99	35.60	118
<b>Recovery rate in total [%]</b>					<b>129</b>

Sample (n=5)	Mean value measured in original sample [µg/l]	Spike [µg/l]	Vitamin B <sub>6</sub> expected [µg/l]	Vitamin B <sub>6</sub> measured [µg/l]	Recovery Rate [%]
B	9.46	10	19.46	22.99	135
		20	29.46	32.50	115
<b>Recovery rate in total [%]</b>					<b>125</b>

Sample (n=5)	Mean value measured in original sample [µg/l]	Spike [µg/l]	Vitamin B <sub>6</sub> expected [µg/l]	Vitamin B <sub>6</sub> measured [µg/l]	Recovery Rate [%]
C	8.855	10	18.85	18.31	95
		20	28.85	32.36	118
<b>Recovery rate in total [%]</b>					<b>107</b>

### 13. REFERENCES












1. Ambrosch, A. et al., 2001. Relation between homocysteinaemia and diabetic neuropathy in patients with Type 2 diabetes mellitus. *Diabetic medicine : a journal of the British Diabetic Association*, **18**(3), pp.185–92.
2. Dierkes, J. et al., 2001. Homocysteine lowering effect of different multivitamin preparations in patients with end-stage renal disease. *Journal of renal nutrition : the official journal of the Council on Renal Nutrition of the National Kidney Foundation*, **11**(2), pp.67–72.
3. Dierkes, J. et al., 2001. Vitamin supplementation can markedly reduce the homocysteine elevation induced by fenofibrate. *Atherosclerosis*, **158**(1), pp.161–4.
4. Morris, M.C. et al., 2004. Dietary niacin and the risk of incident Alzheimer's disease and of cognitive decline. *Journal of neurology, neurosurgery, and psychiatry*, **75**(8), pp.1093–9.
5. Selhub, J. et al., 1995. Association between plasma homocysteine concentrations and extracranial carotid-artery stenosis. *The New England journal of medicine*, **332**(5), pp.286–91.
6. Basheer, M.P. et al., 2016. Vitamin B complex and homocysteine status and Cognitive impairment in the elderly among Indian population. *Journal of Neuroscience and Behavioral Health*, **8**(4), pp.20–26.
7. Lindqvist, M., Hellstrom, A. & Henriksson, A.E., 2012. Abdominal aortic aneurysm and the association with serum levels of Homocysteine, vitamins B6, B12 and Folate. *Am J Cardiovasc Dis*, **2**(4), pp.318–322.
8. Loitsch, S.M. et al., 2014. Comparison of a new microbiological assay with a standard high-performance liquid chromatographic method for determination of vitamin B6 in serum. *Clinical laboratory*, **60**(6), pp.1035–41.

### 14. GENERAL NOTES ON THE TEST AND TEST PROCEDURE

- This assay was produced and distributed according to the IVD guidelines of 98/79/EC.
- All reagents in the kit package are for *in vitro* diagnostic use only.
- ID-Vit® is a trademark of Immundiagnostik AG.
- Reagents should not be used beyond the expiration date stated on the kit label.

- Do not interchange different lot numbers of any kit component within the same assay.
- The guidelines for medical laboratories should be followed.
- Incubation time, incubation temperature and pipetting volumes of the components are defined by the producer. Any variation of the test procedure, which is not coordinated with the producer, may influence the results of the test. Immundiagnostik AG can therefore not be held responsible for any damage resulting from incorrect use.
- Warranty claims and complaints regarding deficiencies must be logged within 14 days after receipt of the product. The product should be send to Immundiagnostik AG along with a written complaint.
- Control samples should be analysed with each run.
- The assay should always be performed according to the enclosed manual.

### Used symbols:

	Temperature limitation		Catalogue Number
	In Vitro Diagnostic Medical Device		To be used with
	Manufacturer		Contains sufficient for <n> tests
	Lot number		Use by
	Attention		Consult instructions for use
	Consult specification data sheet		