

ISO 9001/14000 certified

For research use only

Cat. No. IP11176 | 48 **Tests**

Feline herpesvirus Detection Kit

Test for the detection of Feline Herpes Virus (Rhinitis) by one-step PCR



Distribuito in ITALIA da

Li StarFish S.r.l.

Via Cavour, 35

20063 Cernusco S/N (MI)

telefono 02-92150794

fax 02-92157285

info@listarfish.it

www.listarfish.it

User Manual

REV.2.2

7. NOTICE

- This product was designed to detect more than 100 copies of target gene(or gene segment). When the copy number of target present in the test reaction is less than 100, a false-negative(a negative test result when the attribute for which the subject is being tested actually exists in that subject) may occur. Use this product For Research Use Only.
- Do not use any reagent after the expiration date.
- Do not use together with reagents of other products.
- Follow the instructions.
- Take care in handling of specimen to minimize risk of infection.
- The PCR process is covered by patents issued and applicable in certain countries. iNTRON Biotechnology, Inc. does not encourage or support the unauthorized or unlicensed use of the PCR process. Use of this product is recommended for persons that either have a license to perform PCR or are not required to obtain a license.

8. TROUBLE SHOOTING

- ① In the case of difficult to interpret results due to non-specific bands.
 - ☞ Reduce amount of template by 1/10 dilution and reacts again.
- ② Preparation of PCR reaction at room temperature may cause the non-specific band.
- ③ All procedure should be carried out on ice.

9. ORDERING INFORMATION

Product	Catalog No.
Feline herpesvirus Detection Kit	IP11176
Viral Gene-spin™ Viral DNA/RNA Extraction kit	17151
SiZer™ 100 DNA Marker	24073

1. DESCRIPTION

Feline rhinotracheitis virus (feline herpesvirus type 1 or FHV-1) causes acute respiratory illness known as rhinotracheitis (or feline herpesvirus infection). The virus affects domestic and wild cats worldwide. Rhinotracheitis is characterized by respiratory symptoms such as sneezing, nasal discharge, rhinitis (inflammation of the nose), and conjunctivitis (inflammation of the membrane lining the eyelid).

It also affects the reproductive tract and can cause complications during pregnancy. Rhinotracheitis is part of the feline upper respiratory infection complex, which is a group of viral and bacterial infections (e.g., calicivirus, chlamydiosis) that cause sneezing and discharge from the eyes and nose. Cats often have two or more of these upper respiratory infections at the same time, and FHV-1 is one of the most common. FHV-1 can be transmitted by direct contact with an infected cat's mouth, nose, or eye discharge. Several days of close contact are necessary for infection to occur.

Feline herpesvirus Detection Kit is direct detection of swine influenza virus on the basis of a genetic database, so it can diagnose very fast and accurately. It can amplify only specific gene using the PCR (Polymerase Chain Reaction) method, and take only 2~3 hours for detection. Therefore, it is a very fast accurate, reliable technique.

2. STORAGE

The components of **Feline herpesvirus Detection Kit** should be stored at -20°C, under this condition, the kit is stable until expiration date stated on the label.

3. CONTENTS

Feline herpesvirus PCR Pre-mixture	48 tubes
DNase/RNase-free water (white cap)	1 vial
FHV positive control (Yellow cap)	2 vial

Component in 20µl reaction
i-StarTaq™ DNA Polymerase
dNTPs
PCR Reaction buffer
Chemical stabilizer
Gel loading buffer
8-MOP (dissolved in DMSO)
Primers for FHV

4. SPECIMEN

Performs the test with whole blood, nasal, conjunctival or oral swab. The specimen should be stored at -20°C prior to use.

5. ADDITIONAL REQUIRED MATERIALS

- Disposable gloves
- DNA extraction kit (see 6.1 DNA preparation method)
- Pipettes
- Sterile pipette tip
- Vortex mixer
- Centrifuge for microcentrifuge tubes
- Thermal cycler
- Electrophoresis kit
- UV transilluminator

6. PROCEDURE

Please read through the entire procedure before starting.

6.1 DNA Preparation

Various manufacturers offer DNA isolation kits. Please carry out the DNA isolation according to the manufacturer's instructions. The following standard extraction kit is recommended.

Product	Catalog No.	Manufacturer
Viral gene-spin™ Viral DNA/RNA Extraction Kit	17151	iNtRON Biotechnology, Inc.

6.2 Amplification

- ① Prepare appropriate PCR premix tubes and label. And one PCR premix tube for positive control.
- ② Add 2 μ l of template DNA into the PCR premix tube.
- ③ Add 18 μ l of DNase/RNase-free water into the PCR premix tube to total volume as 20 μ l.
- ④ Add 2 μ l of positive control and 18 μ l of RNase-free water into a PCR premix tube for monitoring of amplification and easy interpretation.
- ⑤ Dissolve the blue pellet by pipetting.
Note :The pellet is easily dissolved, by letting the mixture stand at R.T. for 1-2minutes after adding water.
- ⑥ (Optional) Add mineral oil. This step is unnecessary when using a thermal cycler that employs a top heating method (general methods).
- ⑦ Perform PCR reaction of samples as the below process using PCR machine.

PCR cycle		Temp.	Time
1 Cycle	Initial Denaturation	94 °C	5 min.
	Denaturation	94 °C	30 sec.
40 Cycles	Annealing	52 °C	30 sec.
	Extension	72 °C	40 sec.
1 Cycle	Final extension	72 °C	5 min.

6.3 Detection of Amplified Products

- ① Prepare 1.5% agarose gel containing RedSafe™ Nucleic Acid Staining Solution. (Cat. No. 21141)
- ② Load 7 μ l of PCR product and positive control on agarose gel without adding a loading-dye buffer and perform electrophoresis.
- ③ Run electrophoresis by 100V (required about 30~40 minutes).
- ④ Identify the result on ultra-violet (UV) transilluminator.

6.4 Interpretation

- Expected PCR product size : 519 bp




Fig 1. Electrophoresis of PCR product by **Feline herpesvirus Detection Kit**
 Lane M : 100bp Molecular ladder (iNtRON Biotechnology)
 Lane 1~4 : FHV positive sample
 Lane N : Negative control

6.5 Elimination of carry-over contamination

- Each PCR/RT-PCR Pre-mixture contains 8-methoxypsoralen (8-MOP) for preventing of carry-over contamination.
- All PCR products should be discarded after UV irradiation (10 min/365nm) for preventing from carry-over contamination.